



### INSTRUMENT PROCESSING SHEET

 Agency Miccosukee Police Department

 S/N 80-001657

Florida Department of Law Enforcement

 Date In 10/5/2021

 DI Completion Date 10/6/2021
 Ship

 P/U

 H/D

 CMI

 EE

<b>Intake</b> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By <u>DER</u> Date <u>10/6/2021</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)  <b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>DI Temp. Checks</b> By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.09C</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP6286</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP6287</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP6288</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP6286	202010A 10/05/2022																																																											
0.080	MP6287	202010B 10/05/2022																																																											
0.200	MP6288	202010D 10/06/2022																																																											
0.080 DGS	N/A	AG026705 09/23/2022																																																											
<b>Calibration Adjustment</b> <u>DERR 10/7/21</u> By <u>DERR</u> Barometric Pressure Gauge <u>1016</u> ID # <u>68639</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>20420</td> <td>09/22/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> <td>02/05/2023</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202010A</td> <td>10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202010B</td> <td>10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202010D</td> <td>10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705</td> <td>09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	20060	02/10/2022	0.100	MP5098	20420	09/22/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	00521080A2	02/05/2023	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202010A	10/05/2022	0.080	MP6287	202010B	10/05/2022	0.200	MP6288	202010D	10/06/2022	0.080 DGS	N/A	AG026705	09/23/2022	<b>Department Inspection</b> By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP5095	N/A	N/A																																																										
0.040	MP5096	20060	02/10/2022																																																										
0.100	MP5098	20420	09/22/2022																																																										
0.200	MP5100	20160	03/18/2022																																																										
0.300	MP5101	20030	01/21/2022																																																										
0.080 DGS	N/A	00521080A2	02/05/2023																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP6286	202010A	10/05/2022																																																										
0.080	MP6287	202010B	10/05/2022																																																										
0.200	MP6288	202010D	10/06/2022																																																										
0.080 DGS	N/A	AG026705	09/23/2022																																																										
Simulator	Serial Number																																																												
0.000	MP6284																																																												
Interferent	MP6285																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												
Notes/Suggested Service: <u>Optical calibration to bring values closer to nominal. DERR Admin review identified that the barometric pressure was not noted in the cal adjustment. I remember doing it and that the pressure was 1016 just as the Department inspection. DERR 10/7/21.</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  Israel Soto <small>Digitally signed by Israel Soto Date: 2021.10.06 14:01:27 -0400</small> Tech Review / Date _____ Admin Review / Date _____																																																												



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001657, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001657</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MICCOSUKEE PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>10/06/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:51</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/06/2021

Date

  
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MICCOSUKEE PD  
Time of Inspection: 12:51

Date of Inspection: 10/06/2021

Serial Number: 80-001657  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.079	0.197	0.079
0.000	0.048	0.078	0.198	0.079

Standard Deviations	0.0000	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera DAVID E REYES-RIVERA  
Signature and Printed Name

10/06/2021  
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001657	Micosukee Police Department	10/6/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p><b>0.047 to 0.053</b> <input checked="" type="checkbox"/></p> <p>MICOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:48</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:48	Control Test	0.048	10:49	Air Blank	0.000	10:49	Control Test	0.049	10:50	Air Blank	0.000	10:50	Control Test	0.048	10:51	Air Blank	0.000	10:52	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>MICOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:53	Control Test	0.079	10:53	Air Blank	0.000	10:54	Control Test	0.079	10:55	Air Blank	0.000	10:55	Control Test	0.078	10:56	Air Blank	0.000	10:56	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<p><b>0.194 to 0.206</b> <input checked="" type="checkbox"/></p> <p>MICOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:57	Control Test	0.199	10:58	Air Blank	0.000	10:58	Control Test	0.198	10:59	Air Blank	0.000	11:00	Control Test	0.198	11:00	Air Blank	0.000	11:01	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>MICOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:02	Control Test	0.078	11:02	Air Blank	0.000	11:03	Control Test	0.079	11:03	Air Blank	0.000	11:04	Control Test	0.079	11:04	Air Blank	0.000	11:04	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:48																																																																																																																																																	
Control Test	0.048	10:49																																																																																																																																																	
Air Blank	0.000	10:49																																																																																																																																																	
Control Test	0.049	10:50																																																																																																																																																	
Air Blank	0.000	10:50																																																																																																																																																	
Control Test	0.048	10:51																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.079	10:53																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test	0.079	10:55																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.078	10:56																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test	0.199	10:58																																																																																																																																																	
Air Blank	0.000	10:58																																																																																																																																																	
Control Test	0.198	10:59																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test	0.198	11:00																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.078	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.079	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.079	11:04																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001657	Miccosukee Police Department	10/06/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p><b>0.047 to 0.053</b> <input checked="" type="checkbox"/></p> <p>MICCOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:02	Control Test	0.050	07:03	Air Blank	0.000	07:03	Control Test	0.050	07:04	Air Blank	0.000	07:04	Control Test	0.050	07:05	Air Blank	0.000	07:06	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>MICCOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:08</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:07	Control Test	0.080	07:08	Air Blank	0.000	07:08	Control Test	0.080	07:09	Air Blank	0.000	07:09	Control Test	0.079	07:10	Air Blank	0.000	07:11	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<p><b>0.194 to 0.206</b> <input checked="" type="checkbox"/></p> <p>MICCOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:12</td></tr> <tr><td>Control Test</td><td>0.200</td><td>07:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:13</td></tr> <tr><td>Control Test</td><td>0.199</td><td>07:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr> <tr><td>Control Test</td><td>0.199</td><td>07:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:12	Control Test	0.200	07:12	Air Blank	0.000	07:13	Control Test	0.199	07:14	Air Blank	0.000	07:14	Control Test	0.199	07:15	Air Blank	0.000	07:15	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>MICCOSUKEE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-001657 10/06/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:17</td></tr> <tr><td>Control Test</td><td>0.075</td><td>07:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:18</td></tr> <tr><td>Control Test</td><td>0.076</td><td>07:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:19</td></tr> <tr><td>Control Test</td><td>0.076</td><td>07:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0757</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7630</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:17	Control Test	0.075	07:17	Air Blank	0.000	07:18	Control Test	0.076	07:18	Air Blank	0.000	07:19	Control Test	0.076	07:19	Air Blank	0.000	07:19	Control Test Stats			Average	0.0757		Std Dev	0.0006		Rel Std Dev(%)	0.7630	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.050	07:03																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.050	07:04																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.050	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.080	07:08																																																																																																																																																	
Air Blank	0.000	07:08																																																																																																																																																	
Control Test	0.080	07:09																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.079	07:10																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:12																																																																																																																																																	
Control Test	0.200	07:12																																																																																																																																																	
Air Blank	0.000	07:13																																																																																																																																																	
Control Test	0.199	07:14																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.199	07:15																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2896																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:17																																																																																																																																																	
Control Test	0.075	07:17																																																																																																																																																	
Air Blank	0.000	07:18																																																																																																																																																	
Control Test	0.076	07:18																																																																																																																																																	
Air Blank	0.000	07:19																																																																																																																																																	
Control Test	0.076	07:19																																																																																																																																																	
Air Blank	0.000	07:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0757																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7630																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MICCOSUKEE PD  
Time of Inspection: 06:51

Date of Inspection: 10/06/2021

Serial Number: 80-001657  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
COMPLIANCENOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes-Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

10/06/2021  
Date