



INSTRUMENT PROCESSING SHEET

Agency Osceola County SOS/N 80-003937Florida Department of
Law EnforcementDate In 1/22/2021DI Completion Date 1/28/2021☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By MH	Quality Checks	By MH	Date 01/27/2021	Flow Calibration	By	Date																																																											
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Needs Battery</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>214</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP4864</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP5097</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG003005 01/30/2022</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202010A 10/05/2022	0.080	MP4864	202010B 10/05/2022	0.200	MP5097	202010D 10/06/2022	0.080 DGS	N/A	AG003005 01/30/2022																																																	
Simulator	Serial #	Lot #/Exp																																																																
0.050	MP4863	202010A 10/05/2022																																																																
0.080	MP4864	202010B 10/05/2022																																																																
0.200	MP5097	202010D 10/06/2022																																																																
0.080 DGS	N/A	AG003005 01/30/2022																																																																
			<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By MH</th></tr></thead><tbody><tr><td>☒ Battery Replacement</td><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td>☐ Other</td><td></td></tr></tbody></table>				Maintenance		By MH	☒ Battery Replacement	☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement	☐ Other																																																				
Maintenance		By MH																																																																
☒ Battery Replacement	☐ Dry Gas Regulator Replacement																																																																	
☐ Breath Tube Replacement	☐ Other																																																																	
			<table border="1"><thead><tr><th colspan="2">DI Temp. Checks</th><th>By MH</th></tr></thead><tbody><tr><td>☒ Lab Temp °C</td><td><u>22.07</u></td><td></td></tr><tr><td>External Digital Therm. ID#:</td><td><u>300504</u></td><td></td></tr><tr><td>☒ 34°C +/- .2</td><td>Serial #: <u>MP4863</u></td><td></td></tr><tr><td>☒ 34°C +/- .2</td><td>Serial #: <u>MP4864</u></td><td></td></tr><tr><td>☒ 34°C +/- .2</td><td>Serial #: <u>MP5097</u></td><td></td></tr></tbody></table>				DI Temp. Checks		By MH	☒ Lab Temp °C	<u>22.07</u>		External Digital Therm. ID#:	<u>300504</u>		☒ 34°C +/- .2	Serial #: <u>MP4863</u>		☒ 34°C +/- .2	Serial #: <u>MP4864</u>		☒ 34°C +/- .2	Serial #: <u>MP5097</u>																																											
DI Temp. Checks		By MH																																																																
☒ Lab Temp °C	<u>22.07</u>																																																																	
External Digital Therm. ID#:	<u>300504</u>																																																																	
☒ 34°C +/- .2	Serial #: <u>MP4863</u>																																																																	
☒ 34°C +/- .2	Serial #: <u>MP4864</u>																																																																	
☒ 34°C +/- .2	Serial #: <u>MP5097</u>																																																																	
<table border="1"><thead><tr><th colspan="2">Calibration Adjustment</th><th>By</th></tr></thead><tbody><tr><td colspan="2">Barometric Pressure Gauge ID # _____</td><td></td></tr><tr><td>Simulator</td><td>Serial #</td><td>Lot #</td><td>Expiration</td></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Calibration Adjustment		By	Barometric Pressure Gauge ID # _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th colspan="2">Department Inspection</th><th>By MH</th></tr></thead><tbody><tr><td colspan="2">Barometric Pressure ID# <u>28663</u></td><td></td></tr><tr><td>Gauge <u>1022</u></td><td>Instrument <u>1020</u></td><td></td></tr><tr><td colspan="2">Mouth Alcohol Solution Lot # <u>2020-A</u></td><td></td></tr><tr><td colspan="2">Acetone Stock Solution Lot # <u>2019-A</u></td><td></td></tr><tr><td>Simulator</td><td>Serial Number</td></tr><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1015</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table>		Department Inspection		By MH	Barometric Pressure ID# <u>28663</u>			Gauge <u>1022</u>	Instrument <u>1020</u>		Mouth Alcohol Solution Lot # <u>2020-A</u>			Acetone Stock Solution Lot # <u>2019-A</u>			Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Calibration Adjustment		By																																																																
Barometric Pressure Gauge ID # _____																																																																		
Simulator	Serial #	Lot #	Expiration																																																															
0.000		N/A	N/A																																																															
0.040																																																																		
0.100																																																																		
0.200																																																																		
0.300																																																																		
0.080 DGS	N/A																																																																	
Department Inspection		By MH																																																																
Barometric Pressure ID# <u>28663</u>																																																																		
Gauge <u>1022</u>	Instrument <u>1020</u>																																																																	
Mouth Alcohol Solution Lot # <u>2020-A</u>																																																																		
Acetone Stock Solution Lot # <u>2019-A</u>																																																																		
Simulator	Serial Number																																																																	
0.000	SD1014																																																																	
Interferent	SD1015																																																																	
0.050	MP4863																																																																	
0.080	MP4864																																																																	
0.200	MP5097																																																																	
☐ Post Calibration Adjustment Stability Checks																																																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A																																															
Simulator	Serial #	Lot #	Expiration																																																															
0.050																																																																		
0.080																																																																		
0.200																																																																		
0.080 DGS	N/A																																																																	
Notes/Suggested Service: <u>Clicked dry control instead of stability by mistake.</u>																																																																		
		<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Dry Control</u></td></tr></tbody></table>				Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☒ Other <u>Dry Control</u>																																																			
Attachments																																																																		
☒ Form 41	☐ Post-Stability Checks																																																																	
☒ Stability Checks	☐ Flow Calibration																																																																	
☒ Calibration Certificate	☐ Form 40																																																																	
☐ Calibration Adjustment	☒ Other <u>Dry Control</u>																																																																	
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																
		David Eliezer Reyes-Rivera Digitally signed by David Eliezer Reyes-Rivera Date: 2021.02.09 16:43:37 -0500																																																																
		Tech Review / Date _____ Admin Review / Date _____																																																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 10:57

Date of Inspection: 01/28/2021

Serial Number: 80-003937
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.197	0.080 / 0.081
0.000	0.049	0.078	0.198	0.086 / 0.080
0.000	0.050	0.078	0.198	0.086 / 0.080
0.000	0.050	0.079	0.198	0.085 / 0.080
0.000	0.050	0.080	0.198	0.086 / 0.080
0.000	0.050	0.079	0.198	0.086 / 0.081
0.000	0.050	0.080	0.199	0.086 / 0.080
0.000	0.050	0.080	0.199	0.086 / 0.080
0.000	0.049	0.079	0.199	0.086 / 0.080
0.000	0.051	0.080	0.199	0.086 / 0.079

Standard Deviations	0.0006	0.0007	0.0006	0.0018 / 0.0005
---------------------	--------	--------	--------	-----------------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance UNK CLEARED ROOM MOVED FAN.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

01/28/2021
Date

DERR Digitally signed
by DERR
Date: 2021.02.09
16:42:29 -05'00'

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-003957	Osceola County SO	01/27/2021	MY

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																																
OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr><tr><td>Control Test</td><td>0.049</td><td>14:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr><tr><td>Control Test</td><td>0.049</td><td>14:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr><tr><td>Control Test</td><td>0.050</td><td>14:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	14:13	Control Test	0.049	14:13	Air Blank	0.000	14:14	Control Test	0.049	14:14	Air Blank	0.000	14:15	Control Test	0.050	14:16	Air Blank	0.000	14:16	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	14:17	Control Test	0.079	14:18	Air Blank	0.000	14:19	Control Test	0.078	14:19	Air Blank	0.000	14:20	Control Test	0.079	14:21	Air Blank	0.000	14:21	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr><tr><td>Control Test</td><td>0.198</td><td>14:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1973</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	14:28	Control Test	0.197	14:29	Air Blank	0.000	14:29	Control Test	0.197	14:30	Air Blank	0.000	14:30	Control Test	0.198	14:31	Air Blank	0.000	14:32	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		OSCEOLA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:35</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	14:33	Control Test	0.080	14:33	Air Blank	0.000	14:34	Control Test	0.080	14:34	Air Blank	0.000	14:34	Control Test	0.080	14:35	Air Blank	0.000	14:35	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test	0.049	14:13																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test	0.049	14:14																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.050	14:16																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.079	14:18																																																																																																																																																	
Air Blank	0.000	14:19																																																																																																																																																	
Control Test	0.078	14:19																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test	0.079	14:21																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test	0.197	14:29																																																																																																																																																	
Air Blank	0.000	14:29																																																																																																																																																	
Control Test	0.197	14:30																																																																																																																																																	
Air Blank	0.000	14:30																																																																																																																																																	
Control Test	0.198	14:31																																																																																																																																																	
Air Blank	0.000	14:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:33																																																																																																																																																	
Control Test	0.080	14:33																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																	
Control Test	0.080	14:34																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																	
Control Test	0.080	14:35																																																																																																																																																	
Air Blank	0.000	14:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
MY Operator's Signature	MS Operator's Signature	MS Operator's Signature	MY Operator's Signature																																																																																																																																																

Digitally signed
 by DERR
 Date: 2021.02.09
 16:41:36 -05'00'

OSCEOLA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003937
01/27/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:26
Control Test	0.000	14:26
Air Blank	0.000	14:27

80-803937

Ran Dry Control by Mistake
when turning Stability
Checks. Met Sam hooked up
at time, resulting in the
0.000g/210c Reading.

1/27/2021

A handwritten signature in black ink, appearing to be "M. J. H." or similar, written over the typed name.

(S)
S
D
(1)
(1)
(1)
(1)
(1)
(S)
(S)
(1)
(1)
(1)
(1)





Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003937, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003937</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>OSCEOLA COUNTY S.O.</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/28/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/28/2021

Date


MICHAEL D HAUGHEY,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Digitally signed
by DERR
Date: 2021.02.09
16:39:46 -05'00'

DERR