

INSTRUMENT PROCESSING SHEET

Agency Patrick Air Force Base


S/N 80-005146

Florida Department of
Law Enforcement

Date In 12/9/2021 DI Completion Date 12/13/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By DERR		Quality Checks By DERR Date 12/13/2021		Flow Calibration By DERR Date 12/13/2021																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>135</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.138</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.222</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks		Flow Column # <u>ATP101</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>134</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.228</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																																									
		Maintenance By		☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other																																									
		DI Temp. Checks By DERR		☒ Lab Temp °C <u>20.64C</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP6286</u> ☒ 34°C +/- .2 Serial #: <u>MP6287</u> ☒ 34°C +/- .2 Serial #: <u>MP6288</u>																																									
Calibration Adjustment By DERR		Department Inspection By DERR																																											
Barometric Pressure Gauge <u>1024</u> ID # <u>68639</u>		Barometric Pressure ID# <u>28199</u>																																											
<table><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td>MP5095</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5096</td><td>21070</td><td>03/01/2023</td></tr><tr><td>0.100</td><td>MP5098</td><td>21080</td><td>03/08/2023</td></tr><tr><td>0.200</td><td>MP5100</td><td>20510</td><td>12/03/2022</td></tr><tr><td>0.300</td><td>MP5101</td><td>21030</td><td>02/02/2023</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>00521080A2</td><td>02/05/2023</td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	21070	03/01/2023	0.100	MP5098	21080	03/08/2023	0.200	MP5100	20510	12/03/2022	0.300	MP5101	21030	02/02/2023	0.080 DGS	N/A	00521080A2	02/05/2023	<table><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288		
Simulator	Serial #	Lot #	Expiration																																										
0.000	MP5095	N/A	N/A																																										
0.040	MP5096	21070	03/01/2023																																										
0.100	MP5098	21080	03/08/2023																																										
0.200	MP5100	20510	12/03/2022																																										
0.300	MP5101	21030	02/02/2023																																										
0.080 DGS	N/A	00521080A2	02/05/2023																																										
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☒ Post Calibration Adjustment Stability Checks		Attachments																																											
<table><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td>MP6286</td><td>202010A</td><td>10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B</td><td>10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D</td><td>10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904</td><td>06/08/2023</td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202010A	10/05/2022	0.080	MP6287	202010B	10/05/2022	0.200	MP6288	202010D	10/06/2022	0.080 DGS	N/A	AG115904	06/08/2023	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment		☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other																					
Simulator	Serial #	Lot #	Expiration																																										
0.050	MP6286	202010A	10/05/2022																																										
0.080	MP6287	202010B	10/05/2022																																										
0.200	MP6288	202010D	10/06/2022																																										
0.080 DGS	N/A	AG115904	06/08/2023																																										
Notes/Suggested Service: <u>Optical calibration to adjust barometric value. DERR</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
		Taylor Gutschow		15 14:13:55																																									
		Tech Review / Date		Admin Review / Date																																									



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005146, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005146</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PATRICK AFB</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>12/13/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:07</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form: 69 December 2021

Issuing Authority: Alcohol Testing Program

12/13/2021

Date

DAVID E REYES-RIVERA,

Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PATRICK AFB
Time of Inspection: 11:07

Date of Inspection: 12/13/2021

Serial Number: 80-005146
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.078

Standard Deviations	0.0004	0.0003	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

12/13/2021
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-005146	Patrick Air Force Base	12/13/2021	DERR <i>dm</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:01 Control Test 0.049 09:01 Air Blank 0.000 09:02 Control Test 0.049 09:03 Air Blank 0.000 09:03 Control Test 0.049 09:04 Air Blank 0.000 09:04 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>dm</i>	PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:07 Control Test 0.079 09:07 Air Blank 0.000 09:08 Control Test 0.079 09:09 Air Blank 0.000 09:09 Control Test 0.079 09:10 Air Blank 0.000 09:10 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>dm</i>	PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:11 Control Test 0.199 09:12 Air Blank 0.000 09:13 Control Test 0.198 09:13 Air Blank 0.000 09:14 Control Test 0.198 09:15 Air Blank 0.000 09:15 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2311 Operator's Signature <i>dm</i>	PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:17 Control Test 0.079 09:17 Air Blank 0.000 09:18 Control Test 0.079 09:18 Air Blank 0.000 09:18 Control Test 0.080 09:19 Air Blank 0.000 09:19 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 Operator's Signature <i>dm</i>

PATRICK AB
Intoxilizer - Alcohol Analyzer
Model 8000
12/13/2021
08:00:40
SN 80-005146

Auto Calibration
Max Power Res Value = 32
Auto Range Res Value = 16

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12485, Sum Io = 12423
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1060 (-0.0180)
Sample #2 = 0.1120 (-0.0080)
Sample #3 = 0.0670 (0.0210)
Sample #4 = 0.0930 (-0.0050)
Avg % Abs = 0.0907 (0.0027)
STD DEV = 0.0226 (0.0159)
REL STD DEV = 24.916 (358.044)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1630 (-0.0090)
Sample #2 = 0.1530 (-0.0010)
Sample #3 = 0.1330 (0.0270)
Sample #4 = 0.1340 (0.0100)
Avg % Abs = 0.1400 (0.0120)
STD DEV = 0.0113 (0.0141)
REL STD DEV = 8.050 (117.556)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12477, Sum Io = 12421
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7960 (-0.0190)
Sample #2 = 0.8450 (-0.0200)
Sample #3 = 0.8160 (0.0160)
Sample #4 = 0.7950 (0.0240)
Avg % Abs = 0.8153 (0.0067)
STD DEV = 0.0263 (0.0234)
REL STD DEV = 3.223 (351.568)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4920 (-0.0090)
Sample #2 = 1.5030 (0.0180)
Sample #3 = 1.4980 (0.0170)
Sample #4 = 1.4700 (0.0350)
Avg % Abs = 1.4870 (0.0233)
STD DEV = 0.0165 (0.0101)
REL STD DEV = 1.111 (43.354)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12473, Sum Io = 12419
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8490 (-0.0030)
Sample #2 = 1.8670 (0.0250)
Sample #3 = 1.8590 (0.0330)
Sample #4 = 1.8800 (0.0120)
Avg % Abs = 1.8687 (0.0233)
STD DEV = 0.0106 (0.0106)
REL STD DEV = 0.567 (45.423)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4680 (-0.0090)
Sample #2 = 3.4460 (0.0500)
Sample #3 = 3.4490 (0.0610)
Sample #4 = 3.4460 (0.0530)
Avg % Abs = 3.4470 (0.0547)
STD DEV = 0.0017 (0.0057)
REL STD DEV = 0.050 (10.402)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12470, Sum Io = 12418
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6230 (-0.0070)
Sample #2 = 3.5890 (0.0320)
Sample #3 = 3.6390 (0.0090)
Sample #4 = 3.6210 (0.0350)
Avg % Abs = 3.6163 (0.0253)
STD DEV = 0.0253 (0.0142)
REL STD DEV = 0.700 (56.149)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.6530 (0.0060)
Sample #2 = 6.5860 (0.0670)
Sample #3 = 6.6090 (0.0740)
Sample #4 = 6.6310 (0.0820)
Avg % Abs = 6.6087 (0.0743)
STD DEV = 0.0225 (0.0075)
REL STD DEV = 0.340 (10.097)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12466, Sum Io = 12416
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3370 (-0.0110)
Sample #2 = 5.3030 (0.0230)
Sample #3 = 5.2610 (0.0490)
Sample #4 = 5.2580 (0.0520)
Avg % Abs = 5.2740 (0.0413)
STD DEV = 0.0252 (0.0159)
REL STD DEV = 0.477 (38.583)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.5610 (0.0030)
Sample #2 = 9.5570 (0.1180)
Sample #3 = 9.5380 (0.1350)
Sample #4 = 9.5610 (0.1360)
Avg % Abs = 9.5520 (0.1297)
STD DEV = 0.0123 (0.0101)
REL STD DEV = 0.129 (7.802)

Optical Calibration	
SN:	80-005146
Agency:	Patrick Air Force Base
Date:	12/13/2021
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>HL</i>

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.091
Std Dev = 0.02 Rel Std Dev = 24.92
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.815
Std Dev = 0.03 Rel Std Dev = 3.22
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.869
Std Dev = 0.01 Rel Std Dev = 0.57
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.616
Std Dev = 0.03 Rel Std Dev = 0.70
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.274
Std Dev = 0.03 Rel Std Dev = 0.48
Zero Order Coef = -231.11
First Order Coef = 2604.60
Second Order Coef = 27.72
Standard Deviation = 20.66599

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.140
Std Dev = 0.01 Rel Std Dev = 8.05
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.487
Std Dev = 0.02 Rel Std Dev = 1.11
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.447
Std Dev = 0.00 Rel Std Dev = 0.05
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.609
Std Dev = 0.02 Rel Std Dev = 0.34
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.552
Std Dev = 0.01 Rel Std Dev = 0.13
Zero Order Coef = -183.54
First Order Coef = 1380.38
Second Order Coef = 13.96
Standard Deviation = 18.01567

***** CHANNEL 1 *****
Sample #1 = 3246.00
Sample #2 = 3147.00
Sample #3 = 3236.00
Sample #4 = 3212.00
Average Result = 3198.3333
STD DEV = 46.0471
REL STD DEV = 1.440
***** CHANNEL 2 *****
Sample #1 = 3280.00
Sample #2 = 3267.00
Sample #3 = 3236.00
Sample #4 = 3266.00
Average Result = 3286.3333
STD DEV = 34.3560
REL STD DEV = 1.045

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1024
3 um H2O Adjust (mg/l*10,000) = 611
9 um H2O Adjust (mg/l*10,000) = 523
***** AUTO CAL PASS *****

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 -0.0001
0.100 0.099 0.0006
0.200 0.201 -0.0006
0.300 0.300 0.0002

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0004
0.200 0.201 -0.0005
0.300 0.300 0.0002

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005146	Patrick Air Force Base	12/13/2021	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 SN 80-005146 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	06:48	Control Test	0.050	06:49	Air Blank	0.000	06:50	Control Test	0.050	06:50	Air Blank	0.000	06:51	Control Test	0.050	06:52	Air Blank	0.000	06:52	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 SN 80-005146 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>16:53</td></tr><tr><td>Control Test</td><td>0.081</td><td>16:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:54</td></tr><tr><td>Control Test</td><td>0.080</td><td>16:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>16:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0005</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7167</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	16:53	Control Test	0.081	16:54	Air Blank	0.000	16:54	Control Test	0.080	16:55	Air Blank	0.000	16:56	Control Test	0.080	16:56	Air Blank	0.000	16:57	Control Test Stats			Average	0.0803		Std Dev	0.0005		Rel Std Dev(%)	0.7167		PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 SN 80-005146 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr><tr><td>Control Test</td><td>0.199</td><td>07:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:04</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:06</td></tr><tr><td>Air Blank</td><td>0.100</td><td>07:06</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>1.0005</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	07:03	Control Test	0.199	07:03	Air Blank	0.000	07:04	Control Test	0.198	07:04	Air Blank	0.000	07:05	Control Test	0.198	07:06	Air Blank	0.100	07:06	Control Test Stats			Average	0.1983		Std Dev	1.0005		Rel Std Dev(%)	0.2911		PATRICK AFB Intoxilyzer - Alcohol Analyzer Model 8000 12/13/2021 Software: 8100.27 SN 80-005146 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	07:09	Control Test	0.079	07:09	Air Blank	0.000	07:10	Control Test	0.078	07:10	Air Blank	0.000	07:10	Control Test	0.078	07:11	Air Blank	0.000	07:11	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.050	06:49																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.050	06:50																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.050	06:52																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:53																																																																																																																																																	
Control Test	0.081	16:54																																																																																																																																																	
Air Blank	0.000	16:54																																																																																																																																																	
Control Test	0.080	16:55																																																																																																																																																	
Air Blank	0.000	16:56																																																																																																																																																	
Control Test	0.080	16:56																																																																																																																																																	
Air Blank	0.000	16:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel Std Dev(%)	0.7167																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.199	07:03																																																																																																																																																	
Air Blank	0.000	07:04																																																																																																																																																	
Control Test	0.198	07:04																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test	0.198	07:06																																																																																																																																																	
Air Blank	0.100	07:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	1.0005																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.079	07:09																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.078	07:10																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.078	07:11																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature																																																																																																																																																

INTOXILYZER 8000
Instrument Initialization
05:58 12/13/2021

PATRICK AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005146
12/13/2021
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 6.926
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.871
3: Rate (Liters/min) = 30
SQRT(Diff) = 21.094
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 682
Rounded Intercept = -656273
Correlation = 0.99834

Flow Calibration	
SN:	80-005146
Agency:	Patrick Air Force Base
Date:	12/13/2021
By:	DERR <i>MLL</i>