

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 14:21

Date of Inspection: 04/28/2021

Serial Number: 80-006559
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.079	0.197	0.081
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.079	0.196	0.081
0.000	0.048	0.082	0.197	0.081
0.000	0.049	0.080	0.197	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.196	0.080


Standard Deviations	0.0005	0.0013	0.0008	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

04/28/2021
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - Post	80-006559	Manatee County SO	04/28/2021	mtf

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

MANATEE COUNTY SO
Intoxilyzer - Alcotrol Analyzer
Model 8000 SN 80-016559
04/28/2021
Software: 8100.27

MANATEE COUNTY SO
Intoxilyzer - Alcotrol Analyzer
Model 8000 SN 80-016559
04/28/2021
Software: 8100.27

MANATEE COUNTY SO
Intoxilyzer - Alcotrol Analyzer
Model 8000 SN 80-016559
04/28/2021
Software: 8100.27

Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	10:30	Air Blank	0.000	11:21
Control Test	0.049	10:30	Control Test	0.080	11:21
Air Blank	0.000	10:31	Air Blank	0.000	11:21
Control Test	0.049	10:32	Control Test	0.080	11:22
Air Blank	0.000	10:32	Air Blank	0.000	11:22
Control Test	0.049	10:33	Control Test	0.080	11:23
Air Blank	0.000	10:33	Air Blank	0.000	11:23

Control Test Stats	Average	Std Dev	Rel Std Dev(%)
Control Test Stats	0.0490	0.0000	0.0000
Average	0.0783	0.0006	0.7370
Std Dev	0.0783	0.0006	0.7370
Rel Std Dev(%)	0.0000	0.0000	0.0000

MANATEE COUNTY SO
Intoxilyzer - Alcotrol Analyzer
Model 8000 SN 80-016559
04/28/2021
Software: 8100.27

MANATEE COUNTY SO
Intoxilyzer - Alcotrol Analyzer
Model 8000 SN 80-016559
04/28/2021
Software: 8100.27

mtf

Operator's Signature

mtf

Operator's Signature

mtf

Operator's Signature

mtf

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000
SN 80-006555
04/28/2021 05:14:26

Auto Calibration
Max Power Res Value = 100
Auto Range Res Value = 72

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12576, Sum Io = 12712

Sample % Abs (% Abs Ref)
Sample #1 = 0.0150 (-0.0150)
Sample #2 = -0.0180 (0.0330)
Sample #3 = -0.0140 (0.0420)
Sample #4 = 0.0320 (0.0390)
Avg % Abs = 0.0000 (0.0380)
STD DEV = 0.0278 (0.0046)
REL STD DEV = 2237534720.000 (12.659)

Sample % Abs (% Abs Ref)
Sample #1 = 0.0930 (0.0010)
Sample #2 = 0.0750 (0.0270)
Sample #3 = 0.1090 (0.0080)
Sample #4 = 0.1090 (0.0150)
Avg % Abs = 0.0977 (0.0170)
STD DEV = 0.0196 (0.0095)
REL STD DEV = 20.099 (56.114)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12566, Sum Io = 12706

Sample % Abs (% Abs Ref)
Sample #1 = 0.6950 (-0.0030)
Sample #2 = 0.7400 (-0.0080)
Sample #3 = 0.7570 (-0.0120)
Sample #4 = 0.6580 (0.0260)
Avg % Abs = 0.7317 (0.0020)
STD DEV = 0.0304 (0.0209)
REL STD DEV = 4.151 (1044.031)

Sample % Abs (% Abs Ref)
Sample #1 = 6.5840 (-0.0020)
Sample #2 = 6.6020 (-0.0050)
Sample #3 = 6.5630 (0.0110)
Sample #4 = 6.5990 (-0.0050)
Avg % Abs = 6.5880 (0.0003)
STD DEV = 0.0217 (0.0092)
REL STD DEV = 0.329 (2771.281)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12553, Sum Io = 12701

Sample % Abs (% Abs Ref)
Sample #1 = 5.0880 (-0.0220)
Sample #2 = 5.0660 (0.0000)
Sample #3 = 5.0710 (0.0080)
Sample #4 = 5.1100 (-0.0120)
Avg % Abs = 5.0823 (-0.0013)
STD DEV = 0.0241 (0.0101)
REL STD DEV = 0.474 (754.984)

Sample % Abs (% Abs Ref)
Sample #1 = 9.4750 (-0.0050)
Sample #2 = 9.4670 (0.0200)
Sample #3 = 9.4780 (0.0140)
Sample #4 = 9.4930 (0.0020)
Avg % Abs = 9.4793 (0.0120)
STD DEV = 0.0131 (0.0092)
REL STD DEV = 0.138 (76.376)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12557, Sum Io = 12701

Sample % Abs (% Abs Ref)
Sample #1 = 3.4540 (-0.0060)
Sample #2 = 3.5240 (-0.0280)
Sample #3 = 3.4580 (0.0060)
Sample #4 = 3.5040 (-0.0110)
Avg % Abs = 3.4953 (-0.0110)
STD DEV = 0.0338 (0.0170)
REL STD DEV = 0.968 (154.545)

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.000
Std Dev = 0.03 Rel Std Dev = 2237534720.000

Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.732
Std Dev = 0.03 Rel Std Dev = 4.15

Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.776
Std Dev = 0.02 Rel Std Dev = 1.38
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.495
Std Dev = 0.03 Rel Std Dev = 0.97
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.082
Std Dev = 0.02 Rel Std Dev = 0.47
Zero Order Coef = 6.59
First Order Coef = 2572.30
Second Order Coef = 46.15
Standard Deviation = 29.298475

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.098
Std Dev = 0.02 Rel Std Dev = 20.10
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.457
Std Dev = 0.02 Rel Std Dev = 1.46
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.435
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.588
Std Dev = 0.02 Rel Std Dev = 0.33
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.479
Std Dev = 0.01 Rel Std Dev = 0.14
Zero Order Coef = -117.05
First Order Coef = 1354.41
Second Order Coef = 17.26
Standard Deviation = 22.612339

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0003
0.100 0.100 0.0005
0.200 0.201 -0.0017
0.300 0.300 0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3646.00
Sample #2 = 3624.00
Sample #3 = 3630.00
Sample #4 = 3605.00
Average Result = 3619.6667
STD DEV = 13.0512
REL STD DEV = 0.361

***** CHANNEL 2
Sample #1 = 3407.00
Sample #2 = 3389.00
Sample #3 = 3403.00
Sample #4 = 3403.00
Average Result = 3398.3333
STD DEV = 8.0829
REL STD DEV = 0.238

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 um H2O Adjust (mg/l * 10,000) = 190
9 um H2O Adjust (mg/l * 10,000) = 411
***** AUTO CAL PASS

Optical Calibration	
SN: 80-6555 80-006559	
Agency: Manatee County	
Date: 4/28/21	
By: MH	

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities - Pre	80-006559	Manatee County SO	04/23/21	MX


0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 04/23/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:26</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:25	Control Test	0.048	09:25	Air Blank	0.000	09:26	Control Test	0.049	09:27	Air Blank	0.000	09:27	Control Test	0.049	09:28	Air Blank	0.000	09:28	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 04/23/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:30</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:30	Control Test	0.078	09:31	Air Blank	0.000	09:31	Control Test	0.078	09:32	Air Blank	0.000	09:32	Control Test	0.078	09:33	Air Blank	0.000	09:33	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 04/23/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:35</td></tr> <tr><td>Control Test</td><td>0.197</td><td>09:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:36</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:38</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:35	Control Test	0.197	09:36	Air Blank	0.000	09:36	Control Test	0.198	09:37	Air Blank	0.000	09:38	Control Test	0.198	09:38	Air Blank	0.000	09:39	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006559 04/23/2021 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	09:41	Control Test	0.080	09:41	Air Blank	0.000	09:42	Control Test	0.080	09:42	Air Blank	0.000	09:42	Control Test	0.080	09:43	Air Blank	0.000	09:43	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:25																																																																																																																																																	
Control Test	0.048	09:25																																																																																																																																																	
Air Blank	0.000	09:26																																																																																																																																																	
Control Test	0.049	09:27																																																																																																																																																	
Air Blank	0.000	09:27																																																																																																																																																	
Control Test	0.049	09:28																																																																																																																																																	
Air Blank	0.000	09:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:30																																																																																																																																																	
Control Test	0.078	09:31																																																																																																																																																	
Air Blank	0.000	09:31																																																																																																																																																	
Control Test	0.078	09:32																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test	0.078	09:33																																																																																																																																																	
Air Blank	0.000	09:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:35																																																																																																																																																	
Control Test	0.197	09:36																																																																																																																																																	
Air Blank	0.000	09:36																																																																																																																																																	
Control Test	0.198	09:37																																																																																																																																																	
Air Blank	0.000	09:38																																																																																																																																																	
Control Test	0.198	09:38																																																																																																																																																	
Air Blank	0.000	09:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.080	09:41																																																																																																																																																	
Air Blank	0.000	09:42																																																																																																																																																	
Control Test	0.080	09:42																																																																																																																																																	
Air Blank	0.000	09:42																																																																																																																																																	
Control Test	0.080	09:43																																																																																																																																																	
Air Blank	0.000	09:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

MX
Operator's Signature

MX
Operator's Signature

MX
Operator's Signature

MX
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006559, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006559</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>04/28/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/28/2021

Date


MICHAEL D HAUGHEY,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality