



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department

S/N 80-007084

Florida Department of
Law Enforcement

Date In 8/9/2021

DI Completion Date _____

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: 	Quality Checks By <u>DER</u> Date <u>8/10/2021</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>162</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	202010A 10/05/2022	0.080	SD3968	202010B 10/05/2022	0.200	SD3969	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By <u>DERR</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Internal Printer Paper Repl.</u> DI Temp. Checks By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.09C</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	202010A 10/05/2022																																																											
0.080	SD3968	202010B 10/05/2022																																																											
0.200	SD3969	202010D 10/06/2022																																																											
0.080 DGS	N/A	AG026705 09/23/2022																																																											
Calibration Adjustment By <u>DERR</u> Barometric Pressure Gauge <u>1018/1018</u> ID # <u>68639/68639</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5096</td> <td>20060</td> <td>2/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5098</td> <td>20420</td> <td>09/22/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2</td> <td>02/05/2023</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>202010A</td> <td>10/05/2022</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>202010B</td> <td>10/05/2022</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>202010D</td> <td>10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705</td> <td>09/23/2022</td> </tr> </tbody> </table> Notes/Suggested Service: <u>Optical calibration to bring values closer to nominal. 3rd calibration BP: 1018 Gauge: 68639 Instrument cannot be calibrated and will be sent out for repair. DERR 8/11/2021.</u> 	Simulator	Serial #	Lot #	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5096	20060	2/10/2022	0.100	MP5098	20420	09/22/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	00521080A2	02/05/2023	Simulator	Serial #	Lot #	Expiration	0.050	SD3967	202010A	10/05/2022	0.080	SD3968	202010B	10/05/2022	0.200	SD3969	202010D	10/06/2022	0.080 DGS	N/A	AG026705	09/23/2022	Department Inspection By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> Attachments <div style="display: flex; justify-content: space-between;"> <div> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment </div> <div> <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u> </div> </div> <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between;"> <div>Tech Review / Date _____</div> <div>Admin Review / Date _____</div> </div>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP5095	N/A	N/A																																																										
0.040	MP5096	20060	2/10/2022																																																										
0.100	MP5098	20420	09/22/2022																																																										
0.200	MP5100	20160	03/18/2022																																																										
0.300	MP5101	20030	01/21/2022																																																										
0.080 DGS	N/A	00521080A2	02/05/2023																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	SD3967	202010A	10/05/2022																																																										
0.080	SD3968	202010B	10/05/2022																																																										
0.200	SD3969	202010D	10/06/2022																																																										
0.080 DGS	N/A	AG026705	09/23/2022																																																										
Simulator	Serial Number																																																												
0.000																																																													
Interferent																																																													
0.050																																																													
0.080																																																													
0.200																																																													

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Miami-Dade Police Department

Serial Number: 80-007084

Time of Inspection:

Date of Inspection: 8/11/2021

Software:

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks: I was not able to calibrate the instrument and had to send it out for repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



David E. Reyes-Rivera

Signature and Printed Name

8/11/2021

Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities 3	80-007084	Miami-Dade Police Department	8/11/2021	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:22	Control Test	0.050	10:23	Air Blank	0.000	10:24	Control Test	0.049	10:24	Air Blank	0.000	10:25	Control Test	0.049	10:25	Air Blank	0.000	10:26	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:27	Control Test	0.078	10:28	Air Blank	0.000	10:28	Control Test	0.079	10:29	Air Blank	0.000	10:29	Control Test	0.080	10:30	Air Blank	0.000	10:31	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr><tr><td>Control Test</td><td>0.190</td><td>10:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr><tr><td>Control Test</td><td>0.192</td><td>10:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr><tr><td>Control Test</td><td>0.192</td><td>10:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1913</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.6035</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:32	Control Test	0.190	10:33	Air Blank	0.000	10:33	Control Test	0.192	10:34	Air Blank	0.000	10:34	Control Test	0.192	10:35	Air Blank	0.000	10:36	Control Test Stats			Average	0.1913		Std Dev	0.0012		Rel Std Dev(%)	0.6035		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:37	Control Test	0.080	10:38	Air Blank	0.000	10:38	Control Test	0.079	10:39	Air Blank	0.000	10:39	Control Test	0.081	10:39	Air Blank	0.000	10:40	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.050	10:23																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test	0.049	10:24																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test	0.049	10:25																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.078	10:28																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.079	10:29																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.080	10:30																																																																																																																																																	
Air Blank	0.000	10:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test	0.190	10:33																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.192	10:34																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test	0.192	10:35																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1913																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.6035																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.080	10:38																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test	0.079	10:39																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.081	10:39																																																																																																																																																	
Air Blank	0.000	10:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
08/11/2021 09:38:03

Auto Calibration
Max Power Res Value = 92
Auto Range Res Value = 70

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12623, 9um lo = 13023

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1590 (-0.0150)
Sample #2 = 0.1520 (0.0220)
Sample #3 = 0.0940 (0.0720)
Sample #4 = 0.1410 (0.0720)
Avg % Abs = 0.1290 (0.0553)
STD DEV = 0.0308 (0.0289)
REL STD DEV = 23.880 (52.170)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0030)
Sample #2 = 0.1110 (-0.0020)
Sample #3 = 0.0730 (-0.0040)
Sample #4 = 0.1070 (0.0120)
Avg % Abs = 0.0970 (0.0020)
STD DEV = 0.0209 (0.0087)
REL STD DEV = 21.526 (435.890)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12606, 9um lo = 13018
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8270 (-0.0100)
Sample #2 = 0.8780 (0.0020)
Sample #3 = 0.8310 (0.0250)
Sample #4 = 0.8890 (0.0120)
Avg % Abs = 0.8660 (0.0130)
STD DEV = 0.0308 (0.0115)
REL STD DEV = 3.557 (88.712)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.4580 (0.0010)
Sample #2 = 1.5210 (-0.0060)
Sample #3 = 1.4740 (0.0000)
Sample #4 = 1.5120 (0.0080)
Avg % Abs = 1.5023 (0.0007)
STD DEV = 0.0249 (0.0070)
REL STD DEV = 1.661 (1053.565)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12602, 9um lo = 13017

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9810 (-0.0140)
Sample #2 = 1.9310 (0.0360)
Sample #3 = 1.9830 (0.0160)
Sample #4 = 1.9890 (0.0230)
Avg % Abs = 1.9677 (0.0250)
STD DEV = 0.0319 (0.0101)
REL STD DEV = 1.621 (40.596)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5790 (-0.0070)
Sample #2 = 3.5980 (0.0120)
Sample #3 = 3.6120 (-0.0100)
Sample #4 = 3.6160 (0.0040)
Avg % Abs = 3.6087 (0.0020)
STD DEV = 0.0095 (0.0111)
REL STD DEV = 0.262 (556.776)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12599, 9um lo = 13015

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7980 (-0.0120)
Sample #2 = 3.7970 (0.0120)
Sample #3 = 3.8010 (0.0240)
Sample #4 = 3.7530 (0.0480)
Avg % Abs = 3.7837 (0.0280)
STD DEV = 0.0266 (0.0183)
REL STD DEV = 0.704 (65.465)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.8830 (0.0130)
Sample #2 = 6.9190 (0.0190)
Sample #3 = 6.9180 (0.0140)
Sample #4 = 6.9100 (0.0200)
Avg % Abs = 6.9157 (0.0177)
STD DEV = 0.0049 (0.0032)
REL STD DEV = 0.071 (18.196)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12595, 9um lo = 13014

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.4360 (0.0020)
Sample #2 = 5.5180 (0.0150)
Sample #3 = 5.5140 (0.0190)
Sample #4 = 5.5290 (0.0200)
Avg % Abs = 5.5203 (0.0180)
STD DEV = 0.0078 (0.0026)
REL STD DEV = 0.141 (14.699)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.9370 (-0.0020)
Sample #2 = 9.9850 (0.0550)
Sample #3 = 10.0130 (0.0390)
Sample #4 = 9.9800 (0.0440)
Avg % Abs = 9.9927 (0.0460)
STD DEV = 0.0178 (0.0082)
REL STD DEV = 0.178 (17.794)

Auto Calibration Data

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.129
Std Dev = 0.03 Rel Std Dev = 23.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.866
Std Dev = 0.03 Rel Std Dev = 3.56
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.968
Std Dev = 0.03 Rel Std Dev = 1.62
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.784
Std Dev = 0.03 Rel Std Dev = 0.70
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.520
Std Dev = 0.01 Rel Std Dev = 0.14
Zero Order Coef = -311.86
First Order Coef = 2527.83
Second Order Coef = 20.78
Standard Deviation = 19.485565

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.097
Std Dev = 0.02 Rel Std Dev = 21.53
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.502
Std Dev = 0.02 Rel Std Dev = 1.66
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.609
Std Dev = 0.01 Rel Std Dev = 0.26
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.916
Std Dev = 0.00 Rel Std Dev = 0.07
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.993
Std Dev = 0.02 Rel Std Dev = 0.18
Zero Order Coef = -104.61
First Order Coef = 1296.09
Second Order Coef = 14.32
Standard Deviation = 21.364449

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.039 0.0006
0.100 0.100 0.0001
0.200 0.200 -0.0004
0.300 0.300 0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %XX%
Samples Taken = 4, Discarded = 1
Channel 1 Data:
Sample #1 = 2907.00
Sample #2 = 2798.00
Sample #3 = 2872.00
Sample #4 = 2867.00
Average Result = 2845.6667
STD DEV = 41.3562
REL STD DEV = 1.453

Channel 2 Data:
Sample #1 = 3405.00
Sample #2 = 3396.00
Sample #3 = 3418.00
Sample #4 = 3451.00
Average Result = 3421.6667
STD DEV = 27.6827
REL STD DEV = 0.809
Dry Gas H2O Adjust Results
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 964
9 um H2O Adjust (mg/l*10,000) = 388
Auto Calibration Pass

Optical Calibration 3	
SN:	80-007084
Agency:	Miami-Dade PD
Date:	8/11/2021
Quadratic Fit:	+/- 0.002g/210L
By:	DERR

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0003
0.100 0.100 0.0004
0.200 0.201 -0.0006
0.300 -0.300 0.0002

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities 2	80-007084	Miami-Dade Police Department	8/11/2021	DERR <i>DK</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr><tr><td>Control Test</td><td>0.049</td><td>08:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr><tr><td>Control Test</td><td>0.050</td><td>08:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr></table> <p><i>DK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:51	Control Test	0.049	08:51	Air Blank	0.000	08:52	Control Test	0.050	08:53	Air Blank	0.000	08:53	Control Test	0.050	08:54	Air Blank	0.000	08:54	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr><tr><td>Control Test</td><td>0.080</td><td>08:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:59</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table> <p><i>DK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	08:56	Control Test	0.080	08:56	Air Blank	0.000	08:57	Control Test	0.080	08:57	Air Blank	0.000	08:58	Control Test	0.079	08:59	Air Blank	0.000	08:59	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:00</td></tr><tr><td>Control Test</td><td>0.193</td><td>09:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:01</td></tr><tr><td>Control Test</td><td>0.193</td><td>09:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr><tr><td>Control Test</td><td>0.193</td><td>09:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1930</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p><i>DK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	09:00	Control Test	0.193	09:01	Air Blank	0.000	09:01	Control Test	0.193	09:02	Air Blank	0.000	09:03	Control Test	0.193	09:03	Air Blank	0.000	09:04	Control Test Stats			Average	0.1930		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr><tr><td>Control Test</td><td>0.081</td><td>09:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr><tr><td>Control Test</td><td>0.080</td><td>09:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></table> <p><i>DK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	09:05	Control Test	0.081	09:05	Air Blank	0.000	09:06	Control Test	0.081	09:06	Air Blank	0.000	09:07	Control Test	0.080	09:07	Air Blank	0.000	09:07	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:51																																																																																																																																																	
Control Test	0.049	08:51																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test	0.050	08:53																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.050	08:54																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test	0.080	08:56																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.080	08:57																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test	0.079	08:59																																																																																																																																																	
Air Blank	0.000	08:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:00																																																																																																																																																	
Control Test	0.193	09:01																																																																																																																																																	
Air Blank	0.000	09:01																																																																																																																																																	
Control Test	0.193	09:02																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test	0.193	09:03																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1930																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:05																																																																																																																																																	
Control Test	0.081	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.081	09:06																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test	0.080	09:07																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
08/11/2021 08:03:33

Auto Calibration
Max Power Res Value = 92
Auto Range Res Value = 70

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12647, 9um lo = 13034

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1480 (-0.0140)
Sample #2 = 0.1180 (0.0550)
Sample #3 = 0.0920 (0.0810)
Sample #4 = 0.0920 (0.0970)
Avg % Abs = 0.1007 (0.0777)
STD DEV = 0.0150 (0.0212)
REL STD DEV = 14.912 (27.293)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0760 (0.0160)
Sample #2 = 0.0850 (0.0220)
Sample #3 = 0.0550 (0.0160)
Sample #4 = 0.0610 (0.0380)
Avg % Abs = 0.0670 (0.0253)
STD DEV = 0.0159 (0.0114)
REL STD DEV = 23.693 (44.891)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12631, 9um lo = 13029

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8830 (-0.0130)
Sample #2 = 0.8320 (0.0380)
Sample #3 = 0.8620 (0.0240)
Sample #4 = 0.8800 (0.0380)
Avg % Abs = 0.8580 (0.0333)
STD DEV = 0.0242 (0.0081)
REL STD DEV = 2.826 (21.210)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5190 (0.0100)
Sample #2 = 1.4770 (0.0290)
Sample #3 = 1.5450 (-0.0030)
Sample #4 = 1.5500 (0.0080)
Avg % Abs = 1.5240 (0.0113)
STD DEV = 0.0408 (0.0163)
REL STD DEV = 2.676 (143.456)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12624, 9um lo = 13025

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9760 (-0.0120)
Sample #2 = 1.9200 (0.0380)
Sample #3 = 1.9460 (0.0340)
Sample #4 = 1.9740 (0.0380)
Avg % Abs = 1.9467 (0.0367)
STD DEV = 0.0270 (0.0023)
REL STD DEV = 1.387 (6.298)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5720 (-0.0100)
Sample #2 = 3.5580 (0.0270)
Sample #3 = 3.5600 (0.0250)
Sample #4 = 3.5680 (0.0240)
Avg % Abs = 3.5620 (0.0253)
STD DEV = 0.0053 (0.0015)
REL STD DEV = 0.149 (6.030)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12616, 9um lo = 13024

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.7940 (-0.0170)
Sample #2 = 3.7490 (0.0230)
Sample #3 = 3.7520 (0.0360)
Sample #4 = 3.7500 (0.0460)
Avg % Abs = 3.7503 (0.0350)
STD DEV = 0.0015 (0.0115)
REL STD DEV = 0.041 (32.950)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.8940 (0.0000)
Sample #2 = 6.8970 (0.0280)
Sample #3 = 6.9020 (0.0330)
Sample #4 = 6.8910 (0.0500)
Avg % Abs = 6.8967 (0.0370)
STD DEV = 0.0055 (0.0115)
REL STD DEV = 0.080 (31.169)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %XX%
Samples Taken = 4, Discarded = 1
3um lo = 12610, 9um lo = 13020

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.4580 (-0.0100)
Sample #2 = 5.4410 (0.0320)
Sample #3 = 5.5110 (0.0080)
Sample #4 = 5.4630 (0.0400)
Avg % Abs = 5.4717 (0.0267)
STD DEV = 0.0358 (0.0167)
REL STD DEV = 0.654 (62.450)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.9530 (0.0080)
Sample #2 = 9.9940 (0.0360)
Sample #3 = 9.9890 (0.0230)
Sample #4 = 9.9580 (0.0490)
Avg % Abs = 9.9803 (0.0360)
STD DEV = 0.0195 (0.0130)
REL STD DEV = 0.195 (36.111)

Optical Calibration 2	
SN:	80-007084
Agency:	Miami-Dade PD
Date:	8/11/2021
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>[Signature]</i>

***** AUTO CAL DATA *****
Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.101
Std Dev = 0.02 Rel Std Dev = 14.91
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.858
Std Dev = 0.02 Rel Std Dev = 2.83
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.947
Std Dev = 0.03 Rel Std Dev = 1.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.750
Std Dev = 0.00 Rel Std Dev = 0.04
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.472
Std Dev = 0.04 Rel Std Dev = 0.65
Zero Order Coef = -256.32
First Order Coef = 2514.93
Second Order Coef = 25.92
Standard Deviation = 16.868681
Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.067
Std Dev = 0.02 Rel Std Dev = 23.69
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.524
Std Dev = 0.04 Rel Std Dev = 2.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.562
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.897
Std Dev = 0.01 Rel Std Dev = 0.08
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.980
Std Dev = 0.02 Rel Std Dev = 0.20
Zero Order Coef = -91.32
First Order Coef = 1303.94
Second Order Coef = 13.60
Standard Deviation = 25.035923

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.040 0.040 -0.0003
0.100 0.099 0.0005
0.200 0.200 -0.0003
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.040 0.040 -0.0005
0.100 0.099 0.0008
0.200 0.201 -0.0005
0.300 0.300 0.0002
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %XX%
Samples Taken = 4, Discarded = 1
Channel 1 Data:
Sample #1 = 2931.00
Sample #2 = 2917.00
Sample #3 = 2905.00
Sample #4 = 2948.00
Average Result = 2923.3333
STD DEV = 22.1886
REL STD DEV = 0.759
Channel 2 Data:
Sample #1 = 3464.00
Sample #2 = 3431.00
Sample #3 = 3421.00
Sample #4 = 3413.00
Average Result = 3421.6667
STD DEV = 9.0185
REL STD DEV = 0.264
Dry Gas H2O Adjust Results
Barometric Pressure = 1018
3 um H2O Adjust (mg/l x 10,000) = 886
9 um H2O Adjust (mg/l x 10,000) = 388
**** AUTO CAL PASS

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-007084	Miami-Dade Police Department	8/11/2021	DERR <i>ML</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:44</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:45</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:47</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:48</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></table> <p><i>ML</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:44	Control Test	0.048	07:45	Air Blank	0.000	07:45	Control Test	0.048	07:46	Air Blank	0.000	07:47	Control Test	0.049	07:47	Air Blank	0.000	07:48	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:50</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:51</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:53</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p><i>ML</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:49	Control Test	0.079	07:50	Air Blank	0.000	07:50	Control Test	0.079	07:51	Air Blank	0.000	07:51	Control Test	0.079	07:52	Air Blank	0.000	07:53	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:54</td></tr><tr><td>Control Test</td><td>0.192</td><td>07:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:55</td></tr><tr><td>Control Test</td><td>0.193</td><td>07:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:56</td></tr><tr><td>Control Test</td><td>0.193</td><td>07:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:57</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1927</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2997</td><td></td></tr></table> <p><i>ML</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:54	Control Test	0.192	07:54	Air Blank	0.000	07:55	Control Test	0.193	07:56	Air Blank	0.000	07:56	Control Test	0.193	07:57	Air Blank	0.000	07:57	Control Test Stats			Average	0.1927		Std Dev	0.0006		Rel Std Dev(%)	0.2997		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/11/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr><tr><td>Control Test</td><td>0.080</td><td>07:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:59</td></tr><tr><td>Control Test</td><td>0.079</td><td>08:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:00</td></tr><tr><td>Control Test</td><td>0.081</td><td>08:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:01</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr></table> <p><i>ML</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:58	Control Test	0.080	07:59	Air Blank	0.000	07:59	Control Test	0.079	08:00	Air Blank	0.000	08:00	Control Test	0.081	08:00	Air Blank	0.000	08:01	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:44																																																																																																																																																	
Control Test	0.048	07:45																																																																																																																																																	
Air Blank	0.000	07:45																																																																																																																																																	
Control Test	0.048	07:46																																																																																																																																																	
Air Blank	0.000	07:47																																																																																																																																																	
Control Test	0.049	07:47																																																																																																																																																	
Air Blank	0.000	07:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:49																																																																																																																																																	
Control Test	0.079	07:50																																																																																																																																																	
Air Blank	0.000	07:50																																																																																																																																																	
Control Test	0.079	07:51																																																																																																																																																	
Air Blank	0.000	07:51																																																																																																																																																	
Control Test	0.079	07:52																																																																																																																																																	
Air Blank	0.000	07:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:54																																																																																																																																																	
Control Test	0.192	07:54																																																																																																																																																	
Air Blank	0.000	07:55																																																																																																																																																	
Control Test	0.193	07:56																																																																																																																																																	
Air Blank	0.000	07:56																																																																																																																																																	
Control Test	0.193	07:57																																																																																																																																																	
Air Blank	0.000	07:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1927																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2997																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:58																																																																																																																																																	
Control Test	0.080	07:59																																																																																																																																																	
Air Blank	0.000	07:59																																																																																																																																																	
Control Test	0.079	08:00																																																																																																																																																	
Air Blank	0.000	08:00																																																																																																																																																	
Control Test	0.081	08:00																																																																																																																																																	
Air Blank	0.000	08:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
08/11/2021 07:01:40

Auto Calibration
Max Power Res Value = 91
Auto Range Res Value = 69

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12622, 9um Io = 13023
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (-0.0030)
Sample #2 = 0.0980 (0.0840)
Sample #3 = 0.1440 (0.1090)
Sample #4 = 0.1010 (0.1760)
Avg % Abs = 0.1143 (0.1230)
STD DEV = 0.0257 (0.0476)
REL STD DEV = 22.509 (38.676)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1190 (-0.0120)
Sample #2 = 0.0870 (0.0020)
Sample #3 = 0.1180 (0.0040)
Sample #4 = 0.0730 (0.0620)
Avg % Abs = 0.0927 (0.0227)
STD DEV = 0.0230 (0.0341)
REL STD DEV = 24.851 (150.346)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12593, 9um Io = 13013
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8470 (-0.0060)
Sample #2 = 0.8420 (0.0630)
Sample #3 = 0.8360 (0.0930)
Sample #4 = 0.8840 (0.1000)
Avg % Abs = 0.8540 (0.0853)
STD DEV = 0.0262 (0.0197)
REL STD DEV = 3.062 (23.034)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5290 (-0.0110)
Sample #2 = 1.4930 (0.0430)
Sample #3 = 1.4920 (0.0490)
Sample #4 = 1.5180 (0.0410)
Avg % Abs = 1.5010 (0.0443)
STD DEV = 0.0147 (0.0042)
REL STD DEV = 0.981 (9.391)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12573, 9um Io = 13002
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.9890 (-0.0250)
Sample #2 = 1.9660 (0.0380)
Sample #3 = 1.9330 (0.0630)
Sample #4 = 1.9870 (0.0650)
Avg % Abs = 1.9620 (0.0553)
STD DEV = 0.0272 (0.0150)
REL STD DEV = 1.387 (27.189)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.5360 (-0.0120)
Sample #2 = 3.5690 (0.0170)
Sample #3 = 3.5640 (0.0240)
Sample #4 = 3.6050 (0.0350)
Avg % Abs = 3.5793 (0.0253)
STD DEV = 0.0224 (0.0091)
REL STD DEV = 0.625 (35.818)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12558, 9um Io = 12992
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.7320 (-0.0040)
Sample #2 = 3.7620 (0.0120)
Sample #3 = 3.7570 (0.0570)
Sample #4 = 3.8110 (0.0510)
Avg % Abs = 3.7767 (0.0400)
STD DEV = 0.0298 (0.0244)
REL STD DEV = 0.790 (61.084)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.8320 (0.0000)
Sample #2 = 6.8520 (0.0230)
Sample #3 = 6.8940 (0.0190)
Sample #4 = 6.9260 (0.0030)
Avg % Abs = 6.8907 (0.0150)
STD DEV = 0.0371 (0.0106)
REL STD DEV = 0.539 (70.553)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12547, 9um Io = 12987
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.4830 (-0.0160)
Sample #2 = 5.4430 (0.0480)
Sample #3 = 5.4670 (0.0480)
Sample #4 = 5.4470 (0.1000)
Avg % Abs = 5.4523 (0.0653)
STD DEV = 0.0129 (0.0300)
REL STD DEV = 0.236 (45.952)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.8970 (0.0120)
Sample #2 = 9.9090 (0.0460)
Sample #3 = 9.9580 (0.0280)
Sample #4 = 9.9340 (0.0800)
Avg % Abs = 9.9337 (0.0513)
STD DEV = 0.0245 (0.0264)
REL STD DEV = 0.247 (51.442)

Optical Calibration	
SN:	80-007084
Agency:	Miami-Dade PD
Date:	8/11/2021
Quadratic Fit:	+/- 0.002g/210L
By:	DERR <i>KAL</i>

AUTO CAL DATA
Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.114
Std Dev = 0.03 Rel Std Dev = 22.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.854
Std Dev = 0.03 Rel Std Dev = 3.06
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.962
Std Dev = 0.03 Rel Std Dev = 1.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.777
Std Dev = 0.03 Rel Std Dev = 0.79
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.452
Std Dev = 0.01 Rel Std Dev = 0.24
Zero Order Coef = -253.39
First Order Coef = 2467.01
Second Order Coef = 35.93
Standard Deviation = 38.630856
Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.093
Std Dev = 0.02 Rel Std Dev = 24.85
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.501
Std Dev = 0.01 Rel Std Dev = 0.98
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.579
Std Dev = 0.02 Rel Std Dev = 0.62
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.891
Std Dev = 0.04 Rel Std Dev = 0.54
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.934
Std Dev = 0.02 Rel Std Dev = 0.25
Zero Order Coef = -99.75
First Order Coef = 1299.23
Second Order Coef = 14.85
Standard Deviation = 25.828741

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0004
0.040 0.040 0.0004
0.100 0.100 0.0004
0.200 0.201 -0.0007
0.300 0.300 0.0003
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 2960.00
Sample #2 = 2823.00
Sample #3 = 2830.00
Sample #4 = 2876.00
Average Result = 2843.0000
STD DEV = 28.7924
REL STD DEV = 1.013
Channel 2
Sample #1 = 3407.00
Sample #2 = 3422.00
Sample #3 = 3394.00
Sample #4 = 3388.00
Average Result = 3401.3333
STD DEV = 18.1475
REL STD DEV = 0.534
Dry Gas H2O Adjust Results
Barometric Pressure = 1018
3 um H2O Adjust (mg/l x 10,000) = 966
9 um H2O Adjust (mg/l x 10,000) = 408
AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0006
0.040 0.039 0.0005
0.100 0.099 0.0008
0.200 0.201 -0.0011
0.300 0.300 0.0004

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007084	Miami-Dade Police Department	8/10/2021	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/10/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.047</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.046</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.047</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0467</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2372</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:20	Control Test	0.047	12:21	Air Blank	0.000	12:21	Control Test	0.046	12:22	Air Blank	0.000	12:22	Control Test	0.047	12:23	Air Blank	0.000	12:24	Control Test Stats			Average	0.0467		Std Dev	0.0006		Rel Std Dev(%)	1.2372		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/10/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.076</td><td>12:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.077</td><td>12:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr><tr><td>Control Test</td><td>0.076</td><td>12:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0763</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.076	12:25	Air Blank	0.000	12:26	Control Test	0.077	12:27	Air Blank	0.000	12:27	Control Test	0.076	12:28	Air Blank	0.000	12:28	Control Test Stats			Average	0.0763		Std Dev	0.0006		Rel Std Dev(%)	0.7564		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/10/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr><tr><td>Control Test</td><td>0.187</td><td>12:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr><tr><td>Control Test</td><td>0.188</td><td>12:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr><tr><td>Control Test</td><td>0.190</td><td>12:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1883</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.8111</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:29	Control Test	0.187	12:30	Air Blank	0.000	12:31	Control Test	0.188	12:31	Air Blank	0.000	12:32	Control Test	0.190	12:33	Air Blank	0.000	12:33	Control Test Stats			Average	0.1883		Std Dev	0.0015		Rel Std Dev(%)	0.8111		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/10/2021 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	12:34	Control Test	0.080	12:35	Air Blank	0.000	12:35	Control Test	0.080	12:36	Air Blank	0.000	12:36	Control Test	0.080	12:37	Air Blank	0.000	12:37	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:20																																																																																																																																																	
Control Test	0.047	12:21																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.046	12:22																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.047	12:23																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0467																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2372																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.076	12:25																																																																																																																																																	
Air Blank	0.000	12:26																																																																																																																																																	
Control Test	0.077	12:27																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.076	12:28																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test	0.187	12:30																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test	0.188	12:31																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test	0.190	12:33																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1883																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.8111																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.080	12:35																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.080	12:36																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test	0.080	12:37																																																																																																																																																	
Air Blank	0.000	12:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: David Reyes-Rivera on 8/11/2021

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: I-8000 Serial Number: 80-007084

Bill To Address:

Miami-Dade Police Department

ATTN: Sgt Myrttil

1567 NW 79th Avenue

Miami, Florida 33126

Ship to Address:

Florida Department of Law Enforcement

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument continues to fail post calibration stabilities for the .20 range (calibrated 3 times).

Instrument had a DSP fail before the initial quality check, but later recovered and passed
all diagnostic tests.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sergeant Myrttil

Phone #: (305) 785-3706 Email: u305383@MDPD.com

ATP Contact Name: David Reyes-Rivera ATP Email: DavidReyes@fdle.state.fl.us