



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSOS/N 80-007479

Florida Department of Law Enforcement

Date In 7/8/2021DI Completion Date 7/15/2021 Ship P/U H/D CMI EE

Intake By <u>TDG</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input checked="" type="checkbox"/> 12V DC Cable Notes: <u>Two power cords and a plastic baggie containing a zip tie, male adapter, and female adapter.</u>	Quality Checks By <u>TDG</u> Date <u>7/15/2021</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>172</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	MP5092	202010A 10/05/2022															
0.080	MP5093	202010B 10/05/2022															
0.200	MP5094	202010D 10/06/2022															
0.080 DGS	N/A	AG026705 09/23/2022															
		Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____															
		DI Temp. Checks By <u>TDG</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.99</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5093</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5094</u>															

Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width: 100%;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>TDG</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width: 100%;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>SD3963</td></tr> <tr><td>Interferent</td><td>SD1017</td></tr> <tr><td>0.050</td><td>MP5092</td></tr> <tr><td>0.080</td><td>MP5093</td></tr> <tr><td>0.200</td><td>MP5094</td></tr> </tbody> </table>	Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3963																																																												
Interferent	SD1017																																																												
0.050	MP5092																																																												
0.080	MP5093																																																												
0.200	MP5094																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Form 47</u>																																																													

Notes/Suggested Service: First Stability Test (0.05) did not print due to internal printer not being locked into printing position. All values were within range. Adjusted printer position, repeated the 0.05 Stability Test, and was able to print on internal printer paper. (TDG)

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date <u>Michael D. Haughey</u> 2021.07.15 15:25:10 -04'00'	Admin. Review / Date <u>[Signature]</u> 2021.07.19 11:52:11
---	--



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007479

OWNING AGENCY: Hillsborough County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 7/15/2021

AGENCY INSPECTOR: Roger Skipper

ADDRESS: 1201 Orient Road

CITY, STATE, ZIP: Tampa, FL 33619

TELEPHONE NUMBER: 813-267-7154

FAX NUMBER: n/a

EMAIL ADDRESS (if available): RSkipper@hcsso.tampa.fl.us

For Program Office Use Only:

- Registration Issued
- Instrument Added to Evidentiary Instrument Database
- Instrument Added to Monthly Statistics Database
- Contact Information Added to Instrument Database

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007479	Hillsborough CSO	07/15/2021	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	09:58	Air Blank	0.000	10:08
Control Test	0.050	09:59	Control Test	0.082	10:09
Air Blank	0.000	09:59	Air Blank	0.000	10:09
Control Test	0.050	10:00	Control Test	0.082	10:10
Air Blank	0.000	10:00	Air Blank	0.000	10:10
Control Test	0.050	10:01	Control Test	0.082	10:10
Air Blank	0.000	10:02	Air Blank	0.000	10:11
Control Test Stats			Control Test Stats		
Average	0.0500		Average	0.0820	
Std Dev	0.0000		Std Dev	0.0000	
Rel. Std Dev(%)	0.0000		Rel. Std Dev(%)	0.0000	

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	10:03	Air Blank	0.000	10:12
Control Test	0.081	10:04	Control Test	0.202	10:13
Air Blank	0.000	10:04	Air Blank	0.000	10:13
Control Test	0.081	10:05	Control Test	0.202	10:14
Air Blank	0.000	10:06	Air Blank	0.000	10:14
Control Test	0.081	10:06	Control Test	0.201	10:15
Air Blank	0.000	10:07	Air Blank	0.000	10:16
Control Test Stats			Control Test Stats		
Average	0.0810		Average	0.2017	
Std Dev	0.0000		Std Dev	0.0006	
Rel. Std Dev(%)	0.0000		Rel. Std Dev(%)	0.2863	

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/15/2021
Software: 8100.27

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 12:31

Date of Inspection: 07/15/2021

Serial Number: 80-007479
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.050	0.080	0.202	0.082
0.000	0.050	0.080	0.202	0.082
0.000	0.050	0.081	0.202	0.082
0.000	0.050	0.080	0.201	0.082
0.000	0.051	0.080	0.202	0.082
0.000	0.050	0.080	0.202	0.082
0.000	0.051	0.080	0.202	0.082
0.000	0.051	0.081	0.202	0.081
0.000	0.050	0.081	0.202	0.082
0.000	0.050	0.081	0.202	0.081


Standard Deviations	0.0004	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

TAYLOR D GUTSCHOW

07/15/2021
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007479, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007479</u>	UNCERTAINTY* ±
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>07/15/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:31</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/15/2021

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality