



INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSOS/N 80-007484Florida Department of
Law EnforcementDate In 7/8/2021DI Completion Date 7/19/2021☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ ☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☒ 12V DC Cable Notes: Two power cords and a plastic baggie containing a zip tie, male adapter, and female adapter. (TDG) _____ _____ _____ _____ _____ _____	Quality Checks By TDG _____ Date <u>7/15/2021</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>179</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A 10/05/2022	0.080	MP5093	202010B 10/05/2022	0.200	MP5094	202010D 10/06/2022	0.080 DGS	N/A	AG026705 09/23/2022	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ DI Temp. Checks By TDG _____ ☒ Lab Temp °C <u>21.81</u> External Digital Therm. ID#: <u>300503</u> ☒ 34°C +/- .2 Serial #: <u>MP5092</u> ☒ 34°C +/- .2 Serial #: <u>MP5093</u> ☒ 34°C +/- .2 Serial #: <u>MP5094</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP5092	202010A 10/05/2022																																																											
0.080	MP5093	202010B 10/05/2022																																																											
0.200	MP5094	202010D 10/06/2022																																																											
0.080 DGS	N/A	AG026705 09/23/2022																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Noticed instrument in Disabled Mode so I bypassed the AI to operate. Later realized the instrument was in FACTORY Disabled Mode and brought into Enabled Mode to operate. (TDG)</u> _____ _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By TDG _____ Barometric Pressure ID# <u>28199</u> Gauge <u>1019</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3963</td> </tr> <tr> <td>Interferent</td> <td>SD1017</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 47 / Extra Stabilities</u> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3963																																																												
Interferent	SD1017																																																												
0.050	MP5092																																																												
0.080	MP5093																																																												
0.200	MP5094																																																												



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-007484

OWNING AGENCY: Hillsborough County Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 7/19/2021

AGENCY INSPECTOR: Roger Skipper

ADDRESS: 1201 Orient Road

CITY, STATE, ZIP: Tampa, FL 33619

TELEPHONE NUMBER: 813-267-7154

FAX NUMBER: n/a

EMAIL ADDRESS (if available): RSkipper@hcsso.tampa.fl.us

For Program Office Use Only:

- ☐ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency:

Time of Inspection: 13:37

Date of Inspection: 07/15/2021

Serial Number: 80-007484

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI. COMPLIANCE NOT DETERMINED. Agency name not set at the time. ^{MG} 07/15/2021

Not determined ^{MG} 07/15/2021

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

07/15/2021

Date

HILLSBOROUGH CO SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007484
07/15/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:14
Control Test	0.049	14:15
Air Blank	0.000	14:16
Control Test	0.048	14:16
Air Blank	0.000	14:17
Control Test	0.049	14:18
Air Blank	0.000	14:18
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007484	Hillsborough CSO	07/15/2021	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L #1																																																																																																												
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✗ ≤0.003 of Wet ✗																																																																																																												
Inadvertently printed to external printer. Results all within range. Will attach MG 07/15/2021	HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 07/15/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr></table> MG Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:53	Control Test	0.080	14:54	Air Blank	0.000	14:55	Control Test	0.079	14:55	Air Blank	0.000	14:56	Control Test	0.079	14:57	Air Blank	0.000	14:57	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 07/15/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr><tr><td>Control Test</td><td>0.200</td><td>14:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr><tr><td>Control Test</td><td>0.199</td><td>14:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr><tr><td>Control Test</td><td>0.199</td><td>14:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1993</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr></table> MG Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:22	Control Test	0.200	14:23	Air Blank	0.000	14:23	Control Test	0.199	14:24	Air Blank	0.000	14:24	Control Test	0.199	14:25	Air Blank	0.000	14:26	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		DGS HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 07/15/2021 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr><tr><td>Control Test</td><td>INT*</td><td>14:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr><tr><td>Control Test</td><td>0.083</td><td>14:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0890</td><td></td></tr><tr><td>Std Dev</td><td>0.0131</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>14.6929</td><td></td></tr></table> *Interferent Detect MG Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:27	Control Test	INT*	14:27	Air Blank	0.000	14:28	Control Test	0.083	14:28	Air Blank	0.000	14:29	Control Test	0.080	14:29	Air Blank	0.000	14:29	Control Test Stats			Average	0.0890		Std Dev	0.0131		Rel Std Dev(%)	14.6929	
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:53																																																																																																													
Control Test	0.080	14:54																																																																																																													
Air Blank	0.000	14:55																																																																																																													
Control Test	0.079	14:55																																																																																																													
Air Blank	0.000	14:56																																																																																																													
Control Test	0.079	14:57																																																																																																													
Air Blank	0.000	14:57																																																																																																													
Control Test Stats																																																																																																															
Average	0.0793																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.7277																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:22																																																																																																													
Control Test	0.200	14:23																																																																																																													
Air Blank	0.000	14:23																																																																																																													
Control Test	0.199	14:24																																																																																																													
Air Blank	0.000	14:24																																																																																																													
Control Test	0.199	14:25																																																																																																													
Air Blank	0.000	14:26																																																																																																													
Control Test Stats																																																																																																															
Average	0.1993																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2896																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:27																																																																																																													
Control Test	INT*	14:27																																																																																																													
Air Blank	0.000	14:28																																																																																																													
Control Test	0.083	14:28																																																																																																													
Air Blank	0.000	14:29																																																																																																													
Control Test	0.080	14:29																																																																																																													
Air Blank	0.000	14:29																																																																																																													
Control Test Stats																																																																																																															
Average	0.0890																																																																																																														
Std Dev	0.0131																																																																																																														
Rel Std Dev(%)	14.6929																																																																																																														

Comments: Unknown interferent detected. Allowed air to clear for a few minutes before running an additional 20 stability tests on the DGS, with no interferents detected. Repeated set of stabilities again. All stabilities attached. MG 07/15/2021

DGS

HILLSBOROUGH CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007484
 07/15/2021
 Software: 8100.27

Additional
 Stabilities

07/15/2021

80-007484

Test	g/210L	Time
Air Blank	0.000	14:33
Control Test	0.081	14:34
Air Blank	0.000	14:34
Control Test	0.081	14:34
Air Blank	0.000	14:35
Control Test	0.081	14:35
Air Blank	0.000	14:36
Control Test	0.080	14:36
Air Blank	0.000	14:37
Control Test	0.080	14:37
Air Blank	0.000	14:37
Control Test	0.081	14:38
Air Blank	0.000	14:38
Control Test	0.080	14:39
Air Blank	0.000	14:39
Control Test	0.081	14:39
Air Blank	0.000	14:40
Control Test	0.080	14:40
Air Blank	0.000	14:41
Control Test	0.081	14:41
Air Blank	0.000	14:42
Control Test	0.081	14:42
Air Blank	0.000	14:42
Control Test	0.081	14:43
Air Blank	0.000	14:43
Control Test	0.080	14:44
Air Blank	0.000	14:44
Control Test	0.081	14:44
Air Blank	0.000	14:45
Control Test	0.081	14:45
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Control Test	0.080	14:47
Air Blank	0.000	14:47
Control Test	0.081	14:48
Air Blank	0.000	14:48
Control Test	0.080	14:49
Air Blank	0.000	14:49
Control Test	0.080	14:49
Air Blank	0.000	14:50
Control Test Stats		
Average	0.0806	
Std Dev	0.0005	
Rel Std Dev(%)	0.6337	

Ran 20 additional
 stabilities after
 unknown interferent
 was detected in the
 Stabilities. NO
 interferences detected.
 and all analyses
 in normal range.

TMG 07/15/2021

MG

Operator's Signature

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7484	Hillsborough CSO	07/15/2021	TDG MG

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L #2																																					
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																					
						DGS HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 07/15/2021 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr><tr><td>Control Test</td><td>0.081</td><td>14:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr><tr><td>Control Test</td><td>0.080</td><td>15:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr><tr><td>Control Test</td><td>0.081</td><td>15:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></tbody></table> MG Operator's Signature		Test	g/210L	Time	Air Blank	0.000	14:59	Control Test	0.081	14:59	Air Blank	0.000	14:59	Control Test	0.080	15:00	Air Blank	0.000	15:00	Control Test	0.081	15:01	Air Blank	0.000	15:01	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																									
Air Blank	0.000	14:59																																									
Control Test	0.081	14:59																																									
Air Blank	0.000	14:59																																									
Control Test	0.080	15:00																																									
Air Blank	0.000	15:00																																									
Control Test	0.081	15:01																																									
Air Blank	0.000	15:01																																									
Control Test Stats																																											
Average	0.0807																																										
Std Dev	0.0006																																										
Rel Std Dev(%)	0.7157																																										

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT – INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 11:49

Date of Inspection: 07/19/2021

Serial Number: 80-007484
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#: 202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#: 202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG026705 Exp: 09/23/2022
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.051	0.081	0.199	0.080

Standard Deviations	0.0006	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

07/19/2021
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007484, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007484</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>07/19/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:49</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/19/2021

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1