



Florida Department of
Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Florida Highway Patrol (Troop C) Instrument Serial Number: _____

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☐ Agency Inspection Notification ☒ Agency Contact ☐ Other _____ Date of Notification/Contact: <u>10/08/2021</u> Agency Representative: <u>Travis Donakowski</u>	The following records were audited: ☒ Agency Inspection Reports ☒ Breath Test Instrument Repair/Maintenance Records ☒ Instrument Registration ☒ Dry Gas Standard Certificate(s) of Analysis ☒ Other: <u>Prior Field Notes</u> ☐ Other: _____
Comments: _____ _____ _____		

Instrument/Area:	Equipment:	Supplies:
☒ Clean/Dry ☒ Secure ☒ Limited Access ☐ Other: _____	☒ Proper Number of Simulators ☐ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☒ Class A Glassware	☒ Distilled/Deionized Water ☒ Mouth Alcohol Solution ☒ Acetone Stock Solution ☒ Alcohol Reference Solution ☒ Dry Gas Standard ☒ Mouth Pieces
Comments: <u>Reminded agency to purchase more Alcohol Reference Solution. Only one bottle of each left.</u> _____ _____ _____		

Action:	
☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	Comments: <u>Compliant with Chapter 11D-8, FAC.</u> _____ _____ _____
☐ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

Taylor Dutschow

Signature of Alcohol Testing Program Staff Member

11/05/2021

Date

Original – FDLE

Copy – Agency