



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Daytona Police Department Instrument Serial Number: n/a

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Breath Test Instrument Repair/Maintenance Records, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis, Agency Contact, Date of Notification/Contact: 8/31/2021, Agency Inspector: Justin Dunne, Barometric Pressure: Gauge, Instrument, Instrument Set Up Verified, Other: Other: Other:

Comments: Showed AI how to print registration records from ATP website to add to their records.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Other: Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments:

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use, Comments: Discussed current state of COBRA uploads and best practices. Provided a bottle of acetone and mouth alcohol. Compliant with Chapter 11D-8, FAC.

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.09.02 15:54:23 -04'00' 9/2/2021 Signature of Alcohol Testing Program Staff Member Date

Original - FDLE Copy - Agency