

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Lake Alfred Police Department

Instrument Serial Number: 80-001048

Department Inspection:		Agency Inspection/Agency Contact:	Records Audit:			
	Registration Request for Registration Attached Annual Inspection After Repair Other: Barometric Pressure:	 Agency Inspection Notification Agency Contact Other Date of Notification/Contact: 	 The following records were audited: Agency Inspection Reports Breath Test Instrument Repair/Maintenance Records Instrument Registration Dry Gas Standard Certificate(s) of Analysis 			
	Gauge	6/21/2021				
	Instrument	Agency Inspector: Anthony Gettle	□ Other:			
	Instrument Set Up Verified					
Comments:						
Agency was unable to locate the instrument registration. An ATP copy will be provided for their records.						
Observed AI Gettle perform a complete Agency Inspection from setup to upload completion. The agency						
site and inspection were both compliant with Chapter 11D-8, FAC.						

Instrument/Area:	Equipment:	Supplies:				
Clean/Dry	Proper Number of Simulators	Distilled/Deionized Water				
Secure Secure	Checked Simulators for Air Leak	Mouth Alcohol Solution				
Limited Access	Resistant Seal and Proper	Acetone Stock Solution				
Other:	Temperature	Alcohol Reference Solution				
	Class A Glassware	Dry Gas Standard				
		Mouth Pieces				
Comments:						
Solutions were properly warmed and non-expired. AI Gettle had ATP Form 39 (Agency Inspection						

Procedures) present and clearly visible at all times during the Agency Inspection.

Act	Action:				
	Instrument Complies with Chapter 11D-8, FAC	Comments:			
	Instrument Does Not Comply with Chapter 11D-8, FAC	Recommend sending 10-4D sims out for calibration.			
Remain in Evidentiary Use					
Return to/Place into Evidentiary Use					
	Remove from Evidentiary Use				
Remain Out of Evidentiary Use					
Conduct an Agency Inspection Before Evidentiary Use					
Digitally signed by Taylor Gutschow Date: 2021.06.22 10:24:44 -04'00'		6/21/2021			
Signature of Alcohol Testing Program Staff Member		Date			

Original – FDLE Copy – Agency