Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Nassau Co. SO **Instrument Serial Number:** 80-001053 & 1281

Department Inspection:	Agency Inspection/Agency Contact:	Records Audit:
☐ Registration ☐ Request for Registration Attached ☐ Annual ☐ Inspection After Repair ☐ Other: _____ ☐ Barometric Pressure: _____ Gauge _____ Instrument ☐ Instrument Set Up Verified	☐ Agency Inspection Notification ☑ Agency Contact ☐ Other _____ Date of Notification/Contact: 09/27/21 Agency Inspector: Sgt. Brian Higginbotham, AI	The following records were audited: ☑ Agency Inspection Reports ☑ Breath Test Instrument Repair/Maintenance Records ☑ Instrument Registration ☑ Dry Gas Standard Certificate(s) of Analysis ☐ Other: _____ ☐ Other: _____

Comments:

Records for repair and registration kept with monthly and annual inspection in file cabinet. Other records are kept at State Attorney's office. Agency phone having trouble uploading, dial tone received but getting no connect after multiple tries.

Instrument/Area:	Equipment:	Supplies:
☑ Clean/Dry ☑ Secure ☑ Limited Access ☐ Other:_____	☑ Proper Number of Simulators ☑ Checked Simulators for Air Leak Resistant Seal and Proper Temperature ☑ Class A Glassware	☑ Distilled/Deionized Water ☑ Mouth Alcohol Solution ☑ Acetone Stock Solution ☑ Alcohol Reference Solution ☑ Dry Gas Standard ☑ Mouth Pieces

Comments:

Action:	
☑ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC	Comments:
☑ Remain in Evidentiary Use ☐ Return to/Place into Evidentiary Use ☐ Remove from Evidentiary Use ☐ Remain Out of Evidentiary Use	
☐ Conduct an Agency Inspection Before Evidentiary Use	

Richard A Williams Digitally signed by Richard A Williams Date: 2021.10.04 12:17:22 -04'00'

Signature of Alcohol Testing Program Staff Member **Date**

Original – FDLE **Copy – Agency**