



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Groveland Police Department Instrument Serial Number: 80-001177 and 80-007252

Agency Notification, Records Audit, Agency Inspector/Contact, Date of Notification, Comments: No other records to audit. Agency has no active Intoxilyzer 8000s...

Instrument/Area, Equipment, Supplies. Checklist items: Clean/Dry, Secure, Limited Access, Proper Number of Simulators, Checked Simulators for Air Leak, Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces.

Comments: Dropped off donated items (2 Intoxilyzer 8000s, a printer, Class A glassware, 4 simulators, and a box for shipping), as well as mouth alcohol and acetone. Discussed best practices for uploading and the possibility of Breath Test Instructor training in the future... Recommended fan to promote air flow in the room.

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2022.01.18 13:45:27 -05'00'

1/18/2022

Signature of Alcohol Testing Program Staff Member

Date