



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Lady Lake Police Department Instrument Serial Number: 80-001177

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Agency Contact, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis, Barometric Pressure: Gauge, Instrument, Date of Notification/Contact: 10/12/2021, Agency Inspector: Matt Duryea, Instrument Set Up Verified

Comments: Discussed best practices for uploading. Agency has had the instrument out of service for a while and plans to deregister/donate it soon. Discussed the deregistration/donation process.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Other: Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments: If the agency changes its mind about deregistering 1177, I suggested getting a fan to assist with air flow since the instrument is stored in a small confined space.

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use, Comments: Compliant with Chapter 11D-8, FAC.

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.11.05 10:06:24 -04'00' 11/05/2021 Signature of Alcohol Testing Program Staff Member Date

Original - FDLE Copy - Agency



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Groveland Police Department Instrument Serial Number: 80-001177 and 80-007252

Agency Notification, Records Audit, Agency Inspector/Contact, Date of Notification, Comments: No other records to audit. Agency has no active Intoxilyzer 8000s...

Instrument/Area, Equipment, Supplies. Checklist items: Clean/Dry, Secure, Limited Access, Proper Number of Simulators, Checked Simulators for Air Leak, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces.

Comments: Dropped off donated items (2 Intoxilyzer 8000s, a printer, Class A glassware, 4 simulators, and a box for shipping), as well as mouth alcohol and acetone. Discussed best practices for uploading and the possibility of Breath Test Instructor training in the future... Recommended fan to promote air flow in the room.

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2022.01.18 13:45:27 -05'00'

1/18/2022

Signature of Alcohol Testing Program Staff Member

Date