



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Lake Helen Police Department Instrument Serial Number: 80-001149

Department Inspection: Agency Inspection/Agency Contact: Records Audit: Registration, Agency Inspection Notification, The following records were audited: Agency Inspection Reports, Agency Contact, Instrument Registration, Dry Gas Standard Certificate(s) of Analysis, Other: Simulator Repair Records

Comments: Per Al Mullins, instrument has never been sent to repair, so there were no instrument repair records to locate.

Instrument/Area: Equipment: Supplies: Clean/Dry, Secure, Limited Access, Other: Proper Number of Simulators, Checked Simulators for Air Leak Resistant Seal and Proper Temperature, Class A Glassware, Distilled/Deionized Water, Mouth Alcohol Solution, Acetone Stock Solution, Alcohol Reference Solution, Dry Gas Standard, Mouth Pieces

Comments: Agency did not currently have any deionized water.

Action: Instrument Complies with Chapter 11D-8, FAC, Instrument Does Not Comply with Chapter 11D-8, FAC, Remain in Evidentiary Use, Return to/Place into Evidentiary Use, Remove from Evidentiary Use, Remain Out of Evidentiary Use, Conduct an Agency Inspection Before Evidentiary Use. Comments: Recommended clearly marking expired DGS/ARS to prevent accidental use. Discussed current state of COBRA uploads and best practices. Took Intox with me for its annual inspection.

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.09.02 14:00:18 -04'00' 9/2/2021 Signature of Alcohol Testing Program Staff Member Date

Original - FDLE Copy - Agency