

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:08

Date of Inspection: 03/25/2022

Serial Number: 80-000206

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023 | 0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022 | 0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023 |
|----------------------------|---|---|---|---|
| 0.000 | 0.047 | 0.077 | 0.202 | 0.078 |
| 0.000 | 0.047 | 0.076 | 0.203 | 0.077 |
| 0.000 | 0.047 | 0.076 | 0.203 | 0.077 |

Number of Simulators Used: 4

Remarks:

Int Det:GOT 0.00 READINGS.

Throw out acetone bottle (less than half full)
and open new one, repeated test

AS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ALLISA J STARNES

ALLISA J STARNES

Signature and Printed Name

03/25/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 10:54

Date of Inspection: 04/28/2022

Serial Number: 80-000206
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023 | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|--|---|---|--|
| 0.000 | 0.030 / 0.037 | | | |
| 0.000 | 0.029 / 0.037 | | | |
| 0.000 | 0.029 / 0.037 | | | |

Number of Simulators Used: 4


Remarks:

A F / M A: Improper Sample. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance

Spoke to Taylor
changed out simulator
Worked on second inspection
Attached
AS

The above instrument complies () does not comply (X) with Chapter 11B-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



AILISA J STARNES

Signature and Printed Name

04/28/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: FDLE
Time of Inspection: 10:01

Date of Inspection: 08/29/2022

Serial Number: 80-000206
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: Exp: | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|--|--|--|---|
| | | | | |
| | | | | |
| | | | | |


Number of Simulators Used: 4

Remarks:

A F / M A:GOT READING. Non-compliance: . *Mouth Alcohol test*
1st test - got a reading instead of slope not met
2nd test - Got reading when should of been 0.00. This is due to not waiting long enough after first mouth alcohol test
So I switched out mouth alcohol bottles and restarted whole inspection.
A. Starnes 8/29/22

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALLISA J STARNES

Signature and Printed Name

08/29/2022
Date