



Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Florida Highway Patrol

Instrument Serial Number: 80-000759

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Susan Barge
Date of Inspection: 02-10-2022
Time of Inspection: 16:17
Agency Inspection Discrepancy: [X] Erroneous Information
Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
Lot Number [X] Expiration Date for 0.200g/ 210L [X] Alcohol Reference Solution [ ] Dry Gas Standard is [X] Incorrect [ ] Expired.
FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [ ] REASON for repeating the following test(s); OR the [ ] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[ ] Alcohol Free Subject Test [ ] Mouth Alcohol Test [ ] Alcohol Free Test [ ] Interferent Detect Test
[ ] 0.05 g/210L Test [ ] 0.08 g/210L Test [ ] 0.20 g/210L Test [ ] 0.08 g/210L Dry Gas Standard Test
FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[ ] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[ ] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[ ] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
Other: \_\_\_\_\_

OTHER ELECTRONIC DATA REVIEW
[ ] Login Records Date: \_\_\_\_\_
[ ] Cylinder Change Records Date: \_\_\_\_\_
[ ] Control Test Records Date: \_\_\_\_\_
[ ] Diagnostic Check Records Date: \_\_\_\_\_
Comments: \_\_\_\_\_

CORRECTIVE ACTION
[X] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 03-14-2022 (Date). 04-14-2022
[ ] Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).
[ ] Upload the Agency Inspection(s).
[ ] Remove the instrument from evidentiary use until otherwise directed by the Department.
[ ] No action required
[ ] Other: \_\_\_\_\_

Israel Soto

Digitally signed by Israel Soto
Date: 2022.03.14 10:50:10 -04'00'

Signature of Alcohol Testing Program Staff Member

3/14/2022
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-000759

Time of Inspection: 16:17

Date of Inspection: 02/10/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2023 2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 1482048 Exp: 04/21/2024
0.000	0.048	0.078	0.197	0.081
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SUSAN M BARGE

Signature and Printed Name

02/10/2022

Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-000759

Time of Inspection: 02:30

Date of Inspection: 04/09/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 1482048 Exp: 04/21/2024
0.000	0.050	0.079	0.195	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.079	0.201	0.081


Number of Simulators Used: 5

**Remarks:**

NA

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

04/09/2022

Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 13:29

Date of Inspection: 07/12/2022

Serial Number: 80-000759  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:  
N/A

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NEIL F RUBEL

Signature and Printed Name

07/12/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 12:50

Date of Inspection: 08/11/2022

Serial Number: 80-000759  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024
0.000	0.048	0.077	0.193	0.083
0.000	0.048	0.081	0.199	0.083
0.000	0.048	0.079	0.200	0.083

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

08/11/2022  
Date