AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 13:58

Date of Inspection: 06/03/2022

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
<u> </u>	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1206004 Exp: 09/20/2022
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.078	0.197	0.080

Number	of	Simul	ators	IIsed:	5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

06/03/2022 **Date**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-000784

Time of Inspection: 13:28

Date of Inspection: 07/06/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/22/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1206004 Exp: 09/20/2022
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.079	0.198	0.083
0.000	0.049	0.079	0.198	0.085

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

07/06/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 07:05 Serial Number: 80-000784

Date of Inspection: 08/09/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1206004 Exp: 09/20/2022
0.000	0.047	0.077	0.197	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080

Number	of	Simulators	Used:	5

Remarks:

rks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

08/09/2022 Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Florida Highway Patrol Instrument Serial Number: 80-000784 **AGENCY INSPECTION DATA REVIEW** Agency Inspector: Neil Rubel Date of Inspection: 07-06-2022 Time of Inspection: 13:28 **Agency Inspection Discrepancy:** Incomplete Untimely/Not Received **Erroneous Information** Procedural Other \Box П Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded. Lot Number ⊠ Expiration Date for 0.050g/ 210L ⊠ Alcohol Reference Solution □ Dry Gas Standard is ⊠ Incorrect □ Expired. FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The REASON for repeating the following test(s); OR the Dessible Cause and Corrective Action Taken on the following test(s) was not recorded: ☐ Alcohol Free Subject Test
☐ Mouth Alcohol Test

Alcohol Free Test **Interferent Detect Test** \Box 0.05 g/210L Test 0.08 q/210L Test 0.20 q/210L Test 0.08 g/210L Dry Gas Standard Test FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the П requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use. The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC. Other: OTHER ELECTRONIC DATA REVIEW Comments: **Login Records** Date: **Cylinder Change Records Control Test Records** Date: **Diagnostic Check Records** Date: **CORRECTIVE ACTION** Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by _ Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other:

Israel Soto

Digitally signed by Israel Soto Date: 2022.08.09 10:27:55 -04'00'

Signature of Alcohol Testing Program Staff Member

8/9/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 07:05 Serial Number: 80-000784

Date of Inspection: 08/09/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1206004 Exp: 09/20/2022
0.000	0.047	0.077	0.197	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080

Number	of	Simulators	Used:	5

Remarks:

rks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

08/09/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 00:51

Date of Inspection: 09/19/2022

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402328118 Exp: 01/20/2025
0.000	0.047	0.078	0.197	0.082
0.000	0.047	0.077	0.198	0.082
0.000	0.047	0.077	0.198	0.082

Number of Simulators Used: 5

Remarks: N/A

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

09/19/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 14:01

Date of Inspection: 10/03/2022

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402328118 Exp: 01/20/2025
0.000	0.047	0.078	0.198	0.082
0.000	0.048	0.078	0.200	0.081
0.000	0.048	0.078	0.199	0.081

Number	٥f	Simulators Used: 5	;
namer	OT	Simulators used:	,

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I wold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

10/03/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 08:58

Date of Inspection: 11/08/2022

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402328118 Exp: 01/20/2025
0.000	0.047	0.078	0.198	0.082
0.000	0.047	0.078	0.199	0.082
0.000	0.047	0.078	0.199	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

11/08/2022 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 15:29

Date of Inspection: 12/01/2022

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402328118 Exp: 01/20/2025
0.000	0.047	0.077	0.197	0.082
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.079	0.198	0.082

Number	of	Simulators	IIced.	5	
Mamber	0.7	SIMULACOLS	useu.		

Remarks:

NA

The	above	instrument	complies	(X)	does not	comply	() with	Chapter	11D-8.	FAC.
-----	-------	------------	----------	-----	---	----------	--------	---	--------	---------	--------	------

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SUSAN M BARGE
Signature and Printed Name

12/01/2022 Date