

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Sumter County  
Time of Inspection:

Date of Inspection: 3/2/2022

Serial Number: 08-000816  
Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks: Due to February being a short month and we tend to do our agency inspections closer to the end of the month. The monthly agency inspections wasn't completed until March 2, 2022. The agency inspection will be completed in the middle of every month to prevent this from happening again.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Lt Hollis 604

Lt. Hollis

Signature and Printed Name

03/02/2022

Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER CO SO  
Time of Inspection: 12:32

Date of Inspection: 03/24/2022

Serial Number: 80-000816  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:11821080A2 Exp: 06/05/2023
0.000	0.046	0.076	0.196	0.073 / 0.074
0.000	0.046	0.077	0.196	0.073 / 0.073
0.000	0.046	0.077	0.196	0.073 / 0.073

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. DGS READ BELOW THE ACCEPTABLE RANGE. SPOKE WITH FDLE. *The instrument will be sent to FDLE 7/6/04*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Lt H-604*

TERENCE L HOLLIS

Signature and Printed Name

03/24/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER CO SO  
Time of Inspection: 06:47

Date of Inspection: 08/29/2022

Serial Number: 80-000816  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:213073 Exp: 05/05/2024
0.000	0.045 / 0.046	0.075	0.194	0.078
0.000	0.044 / 0.046	0.075	0.195	0.078
0.000	0.045 / 0.046	0.075	0.194	0.078

Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance.

*Tightened .05 Simulator and tested seal. Then repeated .05 Control test.*

*St. C  
#60246*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*St. C #60246*

PAUL A CAUDLE

Signature and Printed Name

08/29/2022  
Date