



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Umatilla PD (Lake SO)

Instrument Serial Number: 80-000836

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Gerald Corman
Date of Inspection: 05/31/2022
Time of Inspection: 18:56:07
Agency Inspection Discrepancy: [x] Procedural
[x] Lot Number [x] Expiration Date for 0.20 g/ 210L [x] Alcohol Reference Solution [ ] Dry Gas Standard is [ ] Incorrect [x] Expired.

OTHER ELECTRONIC DATA REVIEW
Login Records
Cylinder Change Records
Control Test Records
Diagnostic Check Records
Comments: The 0.20 Alcohol Reference Solution was expired at the time of use. Please mark the inspection as noncompliant and add a remark to the original paperwork that expired solution was used, following the corrective action listed below. Please pull the instrument from evidential use and perform the June 2022 inspection immediately.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 7/20/2022 (Date).

Taylor Dutschow

Signature of Alcohol Testing Program Staff Member

6/20/2022
Date

*\* Amended \**

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD (LAKE SO)  
Time of Inspection: 18:56

Date of Inspection: 05/31/2022

Serial Number: 80-000836  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:  
*mistakenly used expired ref. solution*

*GC 6/22/22 1320 hrs*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*M/P* \_\_\_\_\_ Signature and Printed Name GERALD E CORMAN

05/31/2022  
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Lake CSO (Umatilla PD)

Instrument Serial Number: 80-000836

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Gerald Corman
Date of Inspection: 02/24/2022
Time of Inspection: 09:54:33
Agency Inspection Discrepancy: [X] Erroneous Information
[ ] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[ ] Lot Number [X] Expiration Date for 0.08 g/ 210L [X] Alcohol Reference Solution [ ] Dry Gas Standard is [X] Incorrect [ ] Expired.
[ ] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [ ] REASON for repeating the following test(s); OR the [ ] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[ ] Alcohol Free Subject Test [ ] Mouth Alcohol Test [ ] Alcohol Free Test [ ] Interferent Detect Test
[ ] 0.05 g/210L Test [ ] 0.08 g/210L Test [ ] 0.20 g/210L Test [ ] 0.08 g/210L Dry Gas Standard Test
[ ] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[ ] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[ ] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[ ] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[ ] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[ ] Other: \_\_\_\_\_

OTHER ELECTRONIC DATA REVIEW
[ ] Login Records Date: \_\_\_\_\_
[ ] Cylinder Change Records Date: \_\_\_\_\_
[ ] Control Test Records Date: \_\_\_\_\_
[ ] Diagnostic Check Records Date: \_\_\_\_\_
Comments:
The expiration date (day) for the 0.08 ARS was mistyped.

CORRECTIVE ACTION
[X] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 7/2/2022 (Date).
[ ] Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).
[ ] Upload the Agency Inspection(s).
[ ] Remove the instrument from evidentiary use until otherwise directed by the Department.
[ ] No action required
[ ] Other: \_\_\_\_\_

Taylor Dutschow

Signature of Alcohol Testing Program Staff Member

6/2/2022
Date

Amended

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)  
Time of Inspection: 09:54

Date of Inspection: 02/24/2022

Serial Number: 80-000836  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: <del>07/12/2022</del> 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: TO MUCH MOUTH ALCOHOL.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

m/13 2022 GERALD E CORMAN  
Signature and Printed Name

02/24/2022  
Date