

**(AMENDED)**  
**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: MIAMI BEACH PD  
 Time of Inspection: 13:01

Date of Inspection: 01/26/2022

Serial Number: 80-000879  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A <i>Pat</i> Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1190213 Exp: 08/28/2022
0.000	0.051	0.082	0.204	0.078
0.000	0.051	0.082	0.205	0.078
0.000	0.052	0.082	0.204	0.078

Number of Simulators Used: 5

**Remarks:**

AI K. MILLAN #141 PRESENT

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Pat*

PEDRO A SOCARRAS

Signature and Printed Name

01/26/2022  
Date