

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC

Time of Inspection: 10:16

Date of Inspection: 03/02/2022

Serial Number: 80-000893

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A2 Exp: 05/05/2023
0.000	0.048	0.193 / 0.078	0.197	0.080
0.000	0.049	0.195 / 0.078	0.197	0.079
0.000	0.049	0.196 / 0.078	0.197	0.078

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance WRONG SIMULATOR

. USED THE .02 SIMULATOR RATHER THAN THE

.08 SIMULATOR *Retest was in compliance.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



n965

RICHARD L GILMORE

Signature and Printed Name

03/02/2022

Date