

CR

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC
Time of Inspection: 10:12

Date of Inspection: 03/02/2022

Serial Number: 80-000900
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08121080A2 Exp: 05/05/2023
0.000	0.048	0.076	0.194	0.079
0.000	0.048	0.076	0.194	0.079
0.000	0.049	0.077	0.195	0.080

Number of Simulators Used: 4

Remarks:

Time-Date changed. A F / M A: Range Exceeded.

Range exceeded due to residual mouth alcohol.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 n965

RICHARD L GILMORE

Signature and Printed Name

03/02/2022
Date