

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA  
Time of Inspection: 18:59

Date of Inspection: 02/09/2022

Serial Number: 80-000902  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


**Remarks:**

NOT IN COMPLIANCE Non-compliance: .

INSPECTION DIAGNOSTIC CHECK FAILED, AND INSTRUMENT ABORTED INSPECTION. DAVID REYES WAS CONTACTED REGARDING ISSUE; WITH HIS PERMISSION, I CONDUCTED ANOTHER INSPECTION, WHICH SUBSEQUENTLY PASSED.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAMON J PULASKI

Signature and Printed Name

02/09/2022  
Date