

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O.  
Time of Inspection: 12:05

Date of Inspection: 03/18/2022

Serial Number: 80-000951  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: <del>202108D</del> Exp: <del>08/17/2023</del> 01/12/23	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1474747 Exp: 04/12/2024
/ 0.000	0.048	0.077	0.197	0.080
/ 0.000	0.048	0.078	0.198	0.080
/ 0.000	0.049	0.078	0.198	0.080


Number of Simulators Used: 5

Remarks:

TO CLOSE TO INSTRUMENT AFTER MOUTH ALCOHOL

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAMES R. HESTER

Signature and Printed Name

03/18/2022  
Date