

## Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SATELLITE BEACH P.D.

Serial Number: 80-001001

Time of Inspection:15:08

Date of Inspection: 03/12/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	- X
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023 Lot# 202103D Exp: 08/17/2023	7	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 1497672 Exp: 05/20/2024
0.000	0.051	0.081	0.199	0.079
0.000	0.052	0.081	0.201	0.079
0.000	0.052	0.081	0.200	0.080

Number of Simulators Used: $\underline{4}$	
	w.
Remarks:	and the second s
A F / M A: . Int Det: Ambient Fail.	DISCONNECTED Simulator from instrument, let git the
	Disconnected simulator from instrument, let sit, the Reconnected simulator 19489
The above instrument complies ( X	
I certify that I hold a valid Florid that I performed this inspection in	da Department of Law Enforcement Agency Inspector Permit and accordance with the provisions of Chapter 11D-8, FAC.
5ct (minion	439 DON R TRIEBELL
0	Signature and Printed Name
	03/12/2022

FDLE/ATP Form 40 -- March 2004

Date



## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Satellite Beach Police Department Instrument Serial Number: 80-001001

AGENCY INSPECTION DATA REVIEW					
Agency Inspector: Don Triebell		Date of Inspection: 03-12-20	)22	Time of Inspection: 15:08	
	•	ntimely/Not Received 🗵 her	Erroneo	ous Information	
☐ Agency Inspection Not Conducted or Re	ecords regarding Ag	ency Inspection have not bee	n uploade	d.	
<b>IDENTIFY</b> ■ Lot Number ■ Expiration Date for <u>0.08</u>	<u>80g</u> / 210L ⊠Alcoho	Reference Solution □Dry G	as Standar	d is ⊠Incorrect □Expired.	
<ul> <li>区 FDLE/ATP Form 39 states in part, "If a Remarks section of FDLE/ATP Form 40 test(s); OR the 区 Possible Cause and</li></ul>	Agency Inspection I Corrective Action T	Report – Intoxilyzer 8000. The aken on the following test(s)	e ⊟REAS6 was not re ⊠ ⊔	ON for repeating the following	
□ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." □ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. □ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. □ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.					
☐ The Agency Inspection is noted as "Cor	nplies" when it does	not comply with the requirer	nents of Cl	hapter 11D-8, FAC.	
□ Other:					
OTHER ELECTRONIC DATA REVIEW					
☐ Login Records Date:	Comments:				
Cylinder Change Records Date:					
☐ Control Test Records Date:					
☐ Diagnostic Check Records  Date:					
CORRECTIVE ACTION					
Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by (Date).  Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date).  Upload the Agency Inspection(s).  Remove the instrument from evidentiary use until otherwise directed by the Department.					
☐ No action required ☐ Other:					
Israel Soto  Digitally signed by Israel Soto Date: 2022.04.08 11:07:23 -04'00'					
Signature of Alcohol Testing Program Staff Member  Date					