

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FERNANDINA BEACH PD
Time of Inspection: 15:05

Date of Inspection: 06/08/2022

Serial Number: 80-001059
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023 | 0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:246200803 Exp: 11/05/2022 |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.079 | 0.195 | 0.078 |
| 0.000 | 0.049 | 0.079 | 0.196 | 0.078 |
| 0.000 | 0.049 | 0.079 | 0.197 | 0.078 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JASON E SMITH

Signature and Printed Name

06/08/2022
Date