

↓↓↓↓↓↓↓↓↓↓↓↓↓↓

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ZEPHYRHILLS PD  
Time of Inspection: 09:22

Date of Inspection: 01/31/2022

Serial Number: 80-001076  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.049	0.079	0.201	0.081
0.000	0.050	0.080	0.203	0.081
0.000	0.050	0.081	0.203	0.081


Number of Simulators Used: 4

Remarks:

Modem slot broke  
I retracted in to  
machine

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ALLISA J STARNES

Signature and Printed Name

01/31/2022  
Date