

Agency: Polk CSO

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-001080

AGENCY INSPECTION DATA REVIEW							
Agency Inspector: Cleveland Coke	Time of Inspection: 10:19:28						
Agency Inspection Discrepancy:	omplete Untimely/Not Received	Erroneous Information					
	ocedural 🛛 🛛 Other <u>(Missing Required Rema</u>	<mark>irks)</mark>					
Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.							
□ Lot Number □Expiration Date forg/ 210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired.							
☑ FDLE/ATP Form 39 states in part, "If a te	est must be repeated, the REASON must be entere	d when prompted and recorded in the					
	Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The 🖾 REASON for repeating the following						
	Corrective Action Taken on the following test(s) w						
Alcohol Free Subject Test	Mouth Alcohol Test 🛛 🛛 Alcohol Free Test	Interferent Detect Test					
🗆 0.05 g/210L Test 🛛 🖂	0.08 g/210L Test	0.08 g/210L Dry Gas Standard Test					
□ FDLE/ATP Form 39 states in part, "If the	instrument does not comply with the requirement	ts of Chapter 11D-8, FAC, remove the					
instrument from service and notify the I							
The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency							
Inspection complies with the requirements of Chapter 11D-8, FAC.							
□ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the							
requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of							
Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.							
The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.							
□ Other:							
OTHER ELECTRONIC DATA REVIEW							
Login Records	Comments:						
Date:	The 0.08 ARS Test was repeated. A remark desc	ribing the corrective action taken prior					
Cylinder Change Records	to repeating the test must be included.	• • • • • • • • • •					

Date:
Control Test Records Date:
Diagnostic Check Records

CORRECTIVE ACTION

Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>9/4/2022</u> (Date).

Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).

□ Upload the Agency Inspection(s).

- Remove the instrument from evidentiary use until otherwise directed by the Department.
- No action required
- □ Other: ____

Jaylon Dutschon

Signature of Alcohol Testing Program Staff Member

8/4/2022 Date

FDLE/ATP Form 42 October 2017 Issuing Authority: Alcohol Testing Program PRINTED COPIES UNCONTROLLED For Internal ATP Use Only

Amended

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK CSO Time of Inspection: 10:19

Date of Inspection: 05/11/2022

Serial Number: 80-001080 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
· · · ·	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.048	0.074 / 0.075	0.196	0.078
0.000 *	0.048	0.075 / 0.076	0.197	0.078
0.000	0.048	0.076 / 0.076	0.197	0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

0.08g/2106	Test I tightend the Lid on
Simulator	whited 10 mins and repeated
the test.	Instrument read in An
Acceptable	RAnge.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC. I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I

performed this inspection in accordance with the provisions of Chapter 11D-8, FAC. NO CLEVELAND A COKE 0 Signature and Printed Name 05/11/2022