

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 05:46

Date of Inspection: 03/07/2022

Serial Number: 80-001081  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.073 / 0.071		
0.000	0.048	0.075 / 0.072		
0.000	0.048	0.076 / 0.072		

Number of Simulators Used: 5


**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Problems with a simulator  
 30 3532  
 3/9/22  
 Tightened simulator and allowed to  
 use as normal  
 30 3532  
 6/14/22

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 BARRY NIXON  
 \_\_\_\_\_  
 Signature and Printed Name

03/07/2022  
 Date