

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DELAND P.D.

Time of Inspection: 21:57

Date of Inspection: 07/31/2022

Serial Number: 80-001114

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1433248 Exp: 01/06/2024
0.000	0.046	0.077	0.189 / 0.192	0.079
0.000	0.046	0.077	0.192 / 0.193	0.079
0.000	0.047	0.077	0.192 / 0.194	0.079

Number of Simulators Used: 5


**Remarks:**

20: Control Outside Tolerance.

*Alcohol Solution was not fully to temp.  
Subsequent tests were okay.*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



CRAIG K WALTER

Signature and Printed Name

07/31/2022

Date



Florida Department of  
Law Enforcement

## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Deland Police Department

Instrument Serial Number: 80-001114

### AGENCY INSPECTION DATA REVIEW

Agency Inspector: Craig Walter	Date of Inspection: 07-31-2022	Time of Inspection: 21:57
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input checked="" type="checkbox"/> Procedural <input type="checkbox"/> Other _____		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for _____ g/ 210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input checked="" type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

### OTHER ELECTRONIC DATA REVIEW

<input type="checkbox"/> Login Records Date: _____	Comments: _____
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

### CORRECTIVE ACTION

<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by _____ (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____

Israel Soto

Digitally signed by Israel Soto  
Date: 2022.08.09 10:20:15 -04'00'

Signature of Alcohol Testing Program Staff Member

8/9/2022  
Date