

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW SMYRNA BEACH PD

Serial Number: 80-001152

Time of Inspection: 07:17

Date of Inspection: 06/25/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: _____ Exp: _____
0.000				
0.000				
0.000				

Number of Simulators Used: 4

Remarks:

FORGOT TO ADD INTERFERENT, Purge Fail.

Power Surge During Inspection. WAITED APPROX. 20 min ⁱⁿ * Conducted Inspection AGAIN PASSED
 CR-RAP #1377 6/28/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CR R A P #1377

RALPH A HUNNEFELD

Signature and Printed Name

06/25/2022

Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: New Smyrna Beach Police Department

Instrument Serial Number: 80-001152

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Ralph Hunnefeld
Date of Inspection: 06-25-2022
Time of Inspection: 07:17
Agency Inspection Discrepancy: [X] Procedural
Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
Lot Number [] Expiration Date for ___g/ 210L [] Alcohol Reference Solution [] Dry Gas Standard is [] Incorrect [] Expired.
[X] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [] REASON for repeating the following test(s); OR the [X] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [X] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments: _____

CORRECTIVE ACTION
[X] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by _____ (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Israel Soto

Digitally signed by Israel Soto
Date: 2022.06.28 12:46:46 -04'00'

Signature of Alcohol Testing Program Staff Member

6/28/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW SMYRNA BEACH PD
Time of Inspection: 10:07

Date of Inspection: 08/14/2022

Serial Number: 80-001152
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 33321080A2 Exp: 01/05/2024
0.000	0.047	0.078	0.192	0.080
0.000	0.048	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL R. A. Hunnefeld #1377

RALPH A HUNNEFELD

Signature and Printed Name

08/14/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW SMYRNA BEACH PD
Time of Inspection: 09:05

Date of Inspection: 08/14/2022

Serial Number: 80-001152
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#: 202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 33321080A2 Exp: 01/05/2024
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.080	0.197	0.080
0.000	0.048	0.080	0.198	0.080

Number of Simulators Used: 4

Remarks:

Post Inspection Diagnostic Failure due to ^{pl} pin External printer wire becoming disconnected from instrument

RH

The above instrument complies (~~X~~) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL R A #1777

RALPH A HUNNEFELD

Signature and Printed Name

08/14/2022
Date

NEW SMYRNA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001152
08/14/2022
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

From: [Hunnefeld, Ralph](#)
To: [Soto, Israel](#)
Cc: [Reve, Brooke](#); [Helms, William](#)
Subject: NSBPD Intoxilyzer 80-001152
Date: Sunday, August 14, 2022 10:10:42 AM
Attachments: [image001.png](#)
[scan.pdf](#)
[scan0001.pdf](#)

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Good Morning Mr. Soto,

Today (08/14/2022), I conducted the agency inspection on our instruments. During the diagnostic check (post-inspection) I was informed that the check had failed. I was able to determine while the instrument was conducting the post inspection diagnostic check, the external printer cord became disconnected causing the printer diagnostic to fail. I re-connected the cord and conducted another inspection, which passed with no issues. Also, prior to conducting the second inspection I conducted a separate diagnostics check with passed with no issues. I have attached the Form 40 with the remarks (I forgot to add them when I did the inspection) along with the separate diagnostic check.

Also, that instrument needs to be sent in for the annual department inspection. When would be a good time to send it in and where should it be sent to?

Thank you,

Corporal Ralph A. Hunnefeld #1377
Patrol/Traffic Homicide Investigator

City of New Smyrna Beach
246 Industrial Park Avenue
New Smyrna Beach, FL 32168
Office (386) 424-2250
Non-Emergency (386)424-2220
Emergency 911
rhunnefeld@cityofnsb.com
www.cityofnsb.com



PLEASE NOTE: Florida has a very broad public records law. E-mail communications to or from City of New

Smyrna Beach officials and employees are public records and are available upon request. If you have received this message in error, please notify us immediately by replying to this message and delete it from your computer.

Transmission of sensitive information such as but not limited to social security numbers, credit card information, medical records and criminal justice information (outlined in the [FBI_CJS_Policy](#)) via email is prohibited.

** Please consider the negative environmental impacts associated with printing email **

**VISIT OUR WEBSITE FOR MORE INFORMATION ABOUT THE CITY OF NEW SMYRNA BEACH
FLORIDA**

www.cityofnsb.com