

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PONCE INLET PD
Time of Inspection: 08:33

Date of Inspection: 02/17/2022

Serial Number: 80-001158
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry-Gas Std Test (g/210L) Lot#:1273831 Exp: 01/31/2023
0.000	0.046	0.075	0.194	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.196	0.080

Number of Simulators Used: 4

Remarks: SENT TO FDLE POSS MODEM PROBLEM

UNABLE TO UPLOAD
AI. MIGHT BE MODEM
ISSUE.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SENT ON PI 134

WILLIAM C RYAN

Signature and Printed Name

02/17/2022
Date