



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: **Clermont Police Department**

Instrument Serial Number: **80-001174**

AGENCY INSPECTION DATA REVIEW	
Agency Inspector: Chad Wheeler	Date of Inspection: 09/12/2022 Time of Inspection: 14:02:48
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input checked="" type="checkbox"/> Other (Missing Required Information)	
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.	
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for g/ 210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.	
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input checked="" type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input checked="" type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test	
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.	
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.	
<input type="checkbox"/> Other: _____	

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: Please complete the remark about why the 0.05 ARS Test was repeated and describe what corrective action was taken prior to repeating it.
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION	
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 01/05/2023 (Date) .	
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).	
<input type="checkbox"/> Upload the Agency Inspection(s).	
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.	
<input type="checkbox"/> No action required	
<input type="checkbox"/> Other: _____	

Taylor Dutschow

Signature of Alcohol Testing Program Staff Member

12/5/2022
Date

Amended

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 14:02

Date of Inspection: 09/12/2022

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:02422080A1 Exp: 02/05/2024
0.000	0.040 / 0.050	0.079	0.198	0.079
0.000	0.041 / 0.049	0.080	0.199	0.079
0.000	0.041 / 0.050	0.080	0.199	0.079

Number of Simulators Used: 4

Remarks:

A F / M A:FORGOT MOUTH ALCOHOL. 05: Control Outside Tolerance. IN COMPLIANCE. .05G/210L CONTROL TEST REPEATED DUE TO

Results not within the acceptable range.
- for the second results, I first replaced the .05 solution with a new solution. New solution had the same Lot # and expiration date. Then waited 30 minutes before repeating the control test.

05 # 210 #12/8/22

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Chad Wheeler #210

CHAD T WHEELER

Signature and Printed Name

09/12/2022
Date