

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD
Time of Inspection: 00:40

Date of Inspection: 01/24/2022

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

Remarks:

A F / M A: .

NO MOUTH ALCOHOL USED FOR MA SAMPLE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA

RUSSEL S PENA

Signature and Printed Name

01/24/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD
Time of Inspection: 01:16

Date of Inspection: 01/24/2022

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded.

TOO MUCH MA PRESENT FOR SAMPLE, RESIDUAL FROM 12b

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA 6114

RUSSEL S PENA

Signature and Printed Name

01/24/2022

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: GROVELAND PD
Time of Inspection: 01:50

Date of Inspection: 01/24/2022

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.049	0.079	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.080

Number of Simulators Used: 5

Remarks:
A F / M A:

AF TEST GIVEN SLOPE NOT MET DUE TO RESIDUAL MA FROM PREVIOUS INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA

RUSSEL S PENA

Signature and Printed Name

01/24/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD

Serial Number: 80-001177

Time of Inspection: 01:50

Date of Inspection: 01/24/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.049	0.079	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: .

RESIDUAL MA FROM TOO MUCH PREVIOUS TEST, WAITED TO DISSIPATE PERFORMED AGAIN CORRECTLY.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA

RUSSEL S PENA

Signature and Printed Name

01/24/2022

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: GROVELAND PD
Time of Inspection: 23:26

Date of Inspection: 04/17/2022

Serial Number: 80-001177
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080

Number of Simulators Used: 5

Remarks:

Time-Date changed. **TO MATCH CADS/DISPATCH TIME.**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA

RUSSEL S PENA

Signature and Printed Name

04/17/2022
Date