

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD
Time of Inspection: 16:43

Date of Inspection: 11/27/2022

Serial Number: 80-001205
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. *too much breath alcohol used, retested.*

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #6126 KEITH H KELLEY
Signature and Printed Name

11/27/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD
Time of Inspection: 17:27

Date of Inspection: 11/27/2022

Serial Number: 80-001205
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.048	0.076	0.000 / 0.194	0.079
0.000	0.048	0.077	0.194 / 0.194	0.079
0.000	0.049	0.078	0.194 / 0.195	0.079

Number of Simulators Used: 5


Remarks:

20: Control Outside Tolerance.

Hole in tube, RETESTED AFTER REPLACING TUBE FROM Simulator.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #6126 KEITH H KELLEY
Signature and Printed Name

11/27/2022
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Naval Air Station Pensacola Police Department

Instrument Serial Number: 80-001205

AGENCY INSPECTION DATA REVIEW

Agency Inspector: Keith Kelley

Date of Inspection: 11-27-2022

Time of Inspection: 16:43

Agency Inspection Discrepancy: ☐ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information
☐ Procedural ☐ Other _____

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐ Expiration Date for _____ g/ 210L ☐ Alcohol Reference Solution ☐ Dry Gas Standard is ☐ Incorrect ☐ Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☐ REASON for repeating the following test(s); OR the ☒ Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☒ Alcohol Free Subject Test ☒ Mouth Alcohol Test ☐ Alcohol Free Test ☐ Interferent Detect Test

☐ 0.05 g/210L Test ☐ 0.08 g/210L Test ☒ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.

☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.

☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: _____

OTHER ELECTRONIC DATA REVIEW

☐ Login Records

Date: _____

☐ Cylinder Change Records

Date: _____

☐ Control Test Records

Date: _____

☐ Diagnostic Check Records

Date: _____

Comments:

CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by _____ (Date).

☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).

☐ Upload the Agency Inspection(s).

☐ Remove the instrument from evidentiary use until otherwise directed by the Department.

☐ No action required

☐ Other: _____

Israel Soto

Digitally signed by Israel Soto
Date: 2022.12.07 10:04:11 -05'00'

Signature of Alcohol Testing Program Staff Member

12/7/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD
Time of Inspection: 00:57

Date of Inspection: 12/15/2022

Serial Number: 80-001205
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28021080A2 Exp: 12/05/2023
0.000	0.044 / 0.045	0.050 / 0.077	0.192	0.079
0.000	0.045 / 0.046	0.039 / 0.077	0.192	0.079
0.000	0.045 / 0.046	0.047 / 0.076	0.192	0.079

Number of Simulators Used: 5


Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.

05: RETESTED
08: CHANGED OUT REFERENCE SOLUTION, SAME LOT # AND EXPIRY DATE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

KEITH H KELLEY

12/15/2022
Date