

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection: 09:33

Date of Inspection: 07/11/2022

Serial Number: 80-001309  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG121002 Exp: 07/29/2023
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.196	0.078

Number of Simulators Used: 5

Remarks:

The instrument was in Compliance. Not sure why the alcohol free test shows "NO".

Department inspector Israel Soto has been notified.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brian D. Miller BRIAN D MILLER  
Signature and Printed Name

07/11/2022  
Date