

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.  
Time of Inspection: 16:16

Date of Inspection: 04/29/2022

Serial Number: 80-001345  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211143 Exp: 07/05/2023
0.000	0.047	0.079	0.075 / 0.200	0.081
0.000	0.047	0.080	0.098 / 0.201	0.081
0.000	0.047	0.080	0.120 / 0.200	0.082

Number of Simulators Used: 5


**Remarks:**

20: Control Outside Tolerance.

• 20 test - checked seal, let sit 10 minutes, re-ran test.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

CODY L. WADE

04/29/2022  
Date