

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO  
Time of Inspection: 11:27

03/13/2022  
Date of Inspection: ~~02/13/2022~~ (M)

Serial Number: 80-001348  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14020080A1 Exp: 07/05/2022
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.081	0.199	0.079

Number of Simulators Used: 5

Remarks:  
INSPECTED AT N JAIL

MARCH 2022 INSPECTION, SET DATE WRONG BEFORE INSPECTION (M)

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D.F.P. M. Feltovic #1965

DEP M FELTOVIC

---

Signature and Printed Name

02/13/2022  
Date