

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:01

Date of Inspection: 02/28/2022

Serial Number: 80-001361
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1175968 Exp: 07/30/2022
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.051	0.079	0.200	0.081

Number of Simulators Used: 5


Remarks:

A F / M A: RFI Detect.

In-car body microphone was on at time of inspection. Microphone was turned off and the inspection continued as prompted by instrument

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

WILLIAM D HITCHCOCK

02/28/2022
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 01:44

Date of Inspection: 09/27/2022

Serial Number: 80-001361
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402357111 Exp: 02/10/2025
0.000	0.043 / 0.046	0.079	0.196	0.080
0.000	0.046 / 0.045	0.079	0.197	0.081
0.000	0.045 / 0.046	0.079	0.196	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect.TPR ENTERED ROOM ON RADIO

D BLANKET AROUND SIM .

Control Outside Tolerance: WRAPPED SIM WITH
BLANKET. TEMP RAISED w/11 STANDARD
WTH

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM D HITCHCOCK

Signature and Printed Name

09/27/2022
Date