

Return Material Authorization

Ship to: CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: TJ Graham on 4/19/2022

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001651

Bill To Address:

FDLE FMROC

4700 Terminal Drive

Fort Myers, FL 33991

Ship to Address:

FDLE FMROC

4700 Terminal Drive Suite 1

Ft. Myers, FL 33907

Reason for Return:

DVM monitor shows breath hose temperature and Cell temperature at .0 FDLE is the POC for the repairs.

Please choose one of the following options:

1. I _____, authorize all repairs.
2. I _____, authorize repairs up to \$ _____.
3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: David Reyes-Rivera

Phone #: 239-335-7563 Email: davidreyes@fdle.state.fl.us

ATP Contact Name: David Reyes ATP Email: davidreyes@fdle.state.fl.us