

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 09:36

Date of Inspection: 01/25/2022

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
0.000				
0.000				
0.000				

Number of Simulators Used: 5

**Remarks:**

Int Det: , . Non-compliance: .

CHECKED FOR AIRTIGHT SEAL. COULD NOT FIND ISSUE. DISPOSED OF SOLUTION AFTER SECOND FAIL TEST. REPLACED SOLUTION WITH NEW SOLUTION, MET STANDARD ON SUBSEQUENT TEST

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-3, FAC.



CAMILO J ALMEIDA

Signature and Printed Name

01/25/2022

Date