

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 05:35

Date of Inspection: 03/07/2022

Serial Number: 80-003347  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.047	0.073 / 0.072		
0.000	0.047	0.074 / 0.072		
0.000	0.047	0.072 / 0.073		

Number of Simulators Used: 5

**Remarks:**


08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

*Problem with the simulator  
B-3922  
3/5/22*

*Tighten simulator and allow to  
run for 202 ps 392  
4/14/22*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
BARRY NIXON  
\_\_\_\_\_  
Signature and Printed Name

03/07/2022  
Date