

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:00

Date of Inspection: 12/29/2022

Serial Number: 80-003412
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1433248 Exp: 01/06/2024
0.000	0.051	0.081	0.189 / 0.199	0.081
0.000	0.051	0.081	0.191 / 0.200	0.081
0.000	0.052	0.081	0.186 / 0.200	0.081

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

Re-tightened Heating unit and ^{repeated} test 2nd time. Performed correctly.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gregory S Lamont

GREGORY S LAMONT

Signature and Printed Name

12/29/2022
Date