

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 13:38

Date of Inspection: 01/25/2022

Serial Number: 80-005810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:11821080A2 Exp: 06/05/2023
0.000	INT / 0.049	0.078	0.197	0.081
0.000	/ 0.048	0.079	0.196	0.080
0.000	/ 0.048	0.079	0.196	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect. 05: Interferent Detect. ACCIDENTALLY LEFT SIM SIM WITH INT ATTACHED FOR .05G TEST. REMOVED INT SIM AND REPLACED WITH .05G SIM, MET STANDARD. INSTRUMENT TO CLOSE TO PRINTER WHICH CAUSED RFI. MOVED INSTRUMENT, MET STANDARD.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



CAMILO J ALMEIDA

Signature and Printed Name

01/25/2022
Date