

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 21:38

Date of Inspection: 11/28/2022

Serial Number: 80-006222
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
INT / 0.000				
/ 0.000				
/ 0.000				

Number of Simulators Used: 5

Remarks:

00: Interferent Detect CLEANED AREA. Int Det: Ambient Fail, . Non-compliance:WRONG SOLUTION.

MCS LRI ROOM VENT
MOVED INSTRUMENT TO CLEAN THE AREA
RESTARTED INSPECTION AFTER VENTING AREA

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW KOWALSKI

SANDRA W CHESSHER

Signature and Printed Name

MCS

11/28/2022
Date