

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 13:33

Date of Inspection: 08/30/2022

Serial Number: 80-006234
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1474747 Exp: 04/12/2024
/ 0.000	0.042 / 0.048	0.077	0.200	0.082
/ 0.000	0.043 / 0.048	0.078	0.201	0.083
/ 0.000	0.043 / 0.048	0.078	0.202	0.083

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. SIM DID NOT HAVE PROPER SEAL.

Control outside Tolerance
Mouth Alcohol still in air when I began alcohol Free test.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
Signature and Printed Name

JAMES R HESTER

08/30/2022
Date