

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO  
Time of Inspection: 09:10

Date of Inspection: 07/25/2022

Serial Number: 80-006616  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A3 Exp: 01/05/2024
0.000	0.049	0.079	0.198	0.081
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.078	0.198	0.080

Number of Simulators Used: 5

**Remarks:**

A F / M A: .OUTSIDE TOLERANCE, *Not enough M.A. on first attempt.*

*Amended 8/2/22*  
*Master Deputy Edward Faircloth #533*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*  
EDWARD W FAIRCLOTH  
Signature and Printed Name

07/25/2022  
Date