

Amended

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006632

Time of Inspection: 21:33

Date of Inspection: 02/16/2022

Software: 8100.27


Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 5

**Remarks:**

ACETONE NOT DETECTED ON ACETONE TEST, . AMBIENT FAILURE/ACETONE NOT DETECTED ON ACETONE TEST.

Suspect instrument is unable to collect samples from wet bath simulators based on department inspector field visit on 2/24/22. No bubbles were observed during sample collection of ARS on 2/16 and 2/24. Will send to repair per department inspector  2/24/22

The above instrument complies ( ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM T SMITH

Signature and Printed Name

02/16/2022

Date