

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Florida Highway Patrol  
Time of Inspection: 02/28/2022

Serial Number: 80-006636  
Date of Inspection:

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: \_\_\_\_\_

Remarks:

Instrument is with FDLE/ ATP- Ft Myers for Department Annual Inspection. No Agency inspection was conducted this calendar month

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



William D. Hitchcock

Signature and Printed Name

02/28/2022 \_\_\_\_\_  
Date