

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD

Time of Inspection: 22:09

Date of Inspection: 04/17/2022

Serial Number: 80-007252

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15421080A1 Exp: 08/05/2023
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

Time-Date changed.

TO MARK CADS/DISPATCH TIME.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*

RUSSEL S PENA

Signature and Printed Name

04/17/2022

Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD

Time of Inspection: 12:02

Date of Inspection: 12/09/2022

Serial Number: 80-007252

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07622080A2 Exp: 05/05/2024
/ 0.000	0.047	0.077	0.197	0.078
/ 0.000	0.049	0.078	0.198	0.078
/ 0.000	0.048	0.078	0.198	0.078

Number of Simulators Used: 5

**Remarks:**

00: Ambient Fail.

AMBIENT FAIL DURING AF TESTING, CHECKED  
ALL COMPONENTS, RETESTED WITH NO FAILS.

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Handwritten Signature]*

RUSSEL S PENA

Signature and Printed Name

12/09/2022  
Date



Florida Department of  
Law Enforcement

## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: **Groveland Police Department**

Instrument Serial Number: **80-007252**

### AGENCY INSPECTION DATA REVIEW

Agency Inspector: **Russel Pena**

Date of Inspection: **10/19/2022**

Time of Inspection: **15:53:44**

Agency Inspection Discrepancy: ☐ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information  
☐ Procedural ☒ Other (Missing Required Information)

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐ Expiration Date for g/ 210L ☐ Alcohol Reference Solution ☐ Dry Gas Standard is ☐ Incorrect ☐ Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☒ REASON for repeating the following test(s); OR the ☒ Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☐ Alcohol Free Subject Test ☐ Mouth Alcohol Test ☐ Alcohol Free Test ☐ Interferent Detect Test

☒ 0.05 g/210L Test ☐ 0.08 g/210L Test ☐ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.

☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.

☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: \_\_\_\_\_

### OTHER ELECTRONIC DATA REVIEW

☐ Login Records  
Date: \_\_\_\_\_

☐ Cylinder Change Records  
Date: \_\_\_\_\_

☐ Control Test Records  
Date: \_\_\_\_\_

☐ Diagnostic Check Records  
Date: \_\_\_\_\_

#### Comments:

The Alcohol-Free Test failed due to an Ambient Fail. Please describe the source of the Ambient Fail (if known) and include a remark if the Alcohol Free Test was repeated/failed or if the whole inspection was aborted/restarted.

A second Agency Inspection was subsequently performed and successfully completed for October 2022.

### CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 01/05/2023 (Date).

☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).

☐ Upload the Agency Inspection(s).

☐ Remove the instrument from evidentiary use until otherwise directed by the Department.

☐ No action required

☐ Other: \_\_\_\_\_

*Taylor Gutierrez*

Signature of Alcohol Testing Program Staff Member

12/5/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD

Time of Inspection: 15:53

Date of Inspection: 10/19/2022

Serial Number: 80-007252

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. AMBIENT FAIL RETEST.

CHECKED ALL SEALS - RETESTED OK

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 G114

RUSSEL P. PENA

Signature and Printed Name

10/19/2022  
Date